

Nebraska Department of Agriculture
 Food Safety and Consumer Protection
 Weights and Measures
 P.O. Box 94757
 Lincoln, NE 68509
 402-471-3422

LP GAS PLACE IN SERVICE REPORT

By Registered Service Company

FIRM NUMBER	
TECHNICIAN PHONE	TECHNICIAN ID
TECHNICIAN EMAIL	

MO.	DAY	YEAR	BUSINESS NAME	STREET ADDRESS	CITY	STATE	ZIP CODE

DEVICE NUMBER	UNIT	DISP CODE	CALCULATIONS AND REMARKS: APPROVED () REJECTED () ACTION PENDING () CONDEMNED ()	EMAIL	PHONE
---------------	------	-----------	--	-------	-------

TRUCK ()	PROVER	CERT. DATE: _____	SECURITY YES ()
STATIONARY ()	SERIAL #: _____		SEAL INTACT: NO ()

AUTOMATIC COMPENSATOR – TEST RESULTS

MANUFACTURER	SERIAL NO.	MODEL/SIZE
--------------	------------	------------

TEMPERATURE AT PROVER	_____	_____	_____
KIND OF TEST	NORMAL () SPECIAL ()	NORMAL () SPECIAL ()	NORMAL () SPECIAL ()
PROVER READING (END OF RUN)	TANK _____ PROVER _____	TANK _____ PROVER _____	TANK _____ PROVER _____
PRESSURE			
PRESSURE CORRECTION			
CORRECTED AS TO PRESSURE	_____	_____	_____
LCST CORRECTION			
CORRECTED PROVER READING	_____	_____	_____
CORRECTION FACTOR TO 60°			
CORRECTED TO 60 DEGREES	_____	_____	_____
METER READING			
NET METER ERROR			

UNCOMPENSATED – TEST RESULTS

OWNER NAME

TEMPERATURE OF LIQUID:	GPM () FIRST RUN ()	GPM () FIRST RUN ()	GPM () FIRST RUN ()
AT METER	40G _____ 75G _____ AV _____	40G _____ 75G _____ AV _____	40G _____ 75G _____ AV _____
IN PROVER (END OF RUN)	_____	_____	_____
TEMPERATURE DIFFERENCE	_____	_____	_____
× CORRECTION FACTOR	_____	_____	_____
= TEMPERATURE CORRECTION	_____	_____	_____
KIND OF TEST	NORMAL () SPECIAL ()	NORMAL () SPECIAL ()	NORMAL () SPECIAL ()
P. S. I. G. (END OF RUN)	TANK _____ PROVER _____	TANK _____ PROVER _____	TANK _____ PROVER _____
PROVER READING (END OF RUN)			
PRESSURE CORRECTION	_____	_____	_____
CORRECTED AS TO PRESSURE			
TEMPERATURE CORRECTION	_____	_____	_____
LCST CORRECTION	_____	_____	_____
CORRECTED PROVER READING			
METER READING			
NET METER ERROR			