

Nebraska Standards Laboratory

3721 West Cuming St. Lincoln, NE 68524 (402)-471-2087

Director of Agriculture Steve Wellman PO Box 94947

Lincoln, NE 68509-4947 (402) 471-2341 www.nda.nebraska.gov

Calibration Certificate of Mass

November 4, 2021 Calibration Date:

Certificate Number: 2021-155-1

FSCP Area 10 Submitted By:

Point of Contact: Gene Haase

3721 West Cuming St.

Ph. 402-471-3422

Lincoln, NE 68524

Temp:

email: gene.haase@nebraska.gov

PO Number: N/A

Test Item(s): Cast iron weights

Date Received: November 1, 2021

ID / Asset Number: Area 10 Manufacture: Various

Artifact(s) Description: Serial Number(s): See next page Class Specification: NIST Class F

Material: Cast Iron

Condition: Good (some wear)

Reference Standards Used:

Equipment Used:

Procedure Used: NIST HB 6969, SOP 8 (2019)

Mettler XP 604 Mettler XPR32003

Metrologist: JPL

Environmental Cond.

NSL lb standards

19.4 °C Pressure:

736.35 mmHg

Relative Humidity: 49.2 %

Pertinent Information

• The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. RED print indicates an out-of-compliance reading. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and the uncertainty exceed 95% of the maximum permissible error. All of the tolerances and design specifications (except density, hardness and magnetism) were evaluated according to ASTM E617 (2018) and/or NIST HB 105-1 (2019) for the artifacts designated

- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm³ reference mass density and an air density of 1.2 mg/cm³ at 20 °C.
- It is the end user's responsibility to verify that the weights meet the accuracy requirements outlined in NIST Handbook 44 (2020), Appendix A Fundamental Considerations, when using the weights for calibration of commercial (Legal for Trade) scales.

Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (k), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.



DEPARTMENT OF AGRICULTURE

Nebraska Standards Laboratory

3721 West Cuming St. Lincoln, NE 68524 (402)-471-2087 Director of Agriculture

Steve Wellman P.O. Box 94947 Lincoln, NE 68509-4947 (402) 471-2341 www.nda.nebraska.gov

Calibration Date: November 4, 2021 Certificate Number: 2021-155-1

Calibration Results								
Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)		NIST Class F MPE ± (g)	Assumed Density (g/cm³)
25 lb	D36	-2.06	Y	-0.35	0.14	2	1.1	7.2
25 lb	WM-D51	-0.27	N	-0.27	0.14	2	1.1	7.2
25 lb	WM-D52	-0.30	<u>N</u>	-0.30	0.14	2	1.1	7.2
25 lb	WM-D53	0.07	N	0.07	0.14	2	1.1	7.2
25 lb	WM-D54	1.86	Y	0.72	0.14	2	1.1	7.2
25 lb	WM-D55	-0.03	N N	-0.03	0.14	2	1.1	7.2
25 lb	WM-D56	-0.89	N	-0.89	0.14	2	1.1	7.2
25 lb	WM-D57	0.08	N	0.08	0.14	2	1.1	7.2
25 lb	WM-D58	-0.91	N	-0.91	0.14	2	1.1	7.2
<u>25 lb</u>	WM-D59	-0.85	N	-0.85	0.14	2	1.1	7.2
25 lb	WM-D60	0.12	N	0.12	0.14	2	1.1	7.2
25 lb	WM-D61	-0.09	<u>N</u>	-0.09	0.14	2	1.1	<u>7.2</u>
50 lb	A5C5	2.66	<u>Y</u>	1.20	0.28	2	2.3	7.2
50 lb	<u>C-13</u>	0.06	Ŋ	0.06	0.28	2	2.3	7.2
50 lb	C-C1	-1.98	Y	0.31	0.28	2	2.3	7.2
50 lb	<u>C-C2</u>	2.28	<u>Y</u>	1.30	0.28	2	2.3	7.2
50 lb	C-C3	2.46	Y	0.41	0.28	2	2.3	7.2
50 lb	C-C4	2.92	Y Y	0.81	0.28	2	2.3	7.2
50 lb	C-C6	2.74	Y	1.17	0.28	2	2.3	7.2
50 lb	C-C8	1.10	N Y	1.10	0.28	2	2.3	7.2
50 lb	C-C10	2.59	Y Y	0.86	0.28	2	2.3	7.2
50 lb	C-C14	2.67	Y	0.86	0.28	2	2.3	7.2
50 lb	C-C15	1.54	N Y	1.54	0.28	2	2.3	7.2
50 lb 50 lb	C-C16	2.50 1.20	N Y	0.77 1.20	0.28 0.28	2	2.3 2.3	7.2 7.2
50 lb	C-C18 C-C19	0.62	N N	0.62	0.28	2	2.3	7.2
1000 lb	WME1	96.6	Y	-3.6	<u> </u>	2.019	<u>2.3</u> 45	7.2
1000 lb	WME2	12.5	N T	-3.6 12.5	5.8	2.019	45 45	7.2
1000 lb	WME3	2.5	N N	2.5	<u> </u>	2.019	45 45	7.2
1000 lb	WME4	86.2	IN	-0.7	5.8	2.019	45	7.2
1000 lb	WME5	13.6	N I	13.6	5.8	2.019	45	7.2
1000 lb	WME6	-57.6	Y	9.6	5.8	2.019	45	7.2
1000 lb	WME7	-31.2	Ň	-31.2	5.8	2.019	45	7.2
1000 lb	WME9	- <u>40.0</u>	Y	- <u>51.2</u> -1.4	5.8	2.019	45	7.2
1000 lb	WME10	-40.0 -43.2		4.8	5.8	2.019	45	7.2
1000 lb	WME11	10.0	Ň	10.0	5.8	2.019	45	7.2
1000 lb	WME12	-31.1	N N	-31.1	5.8	2.019	45	7.2
1000 lb	WME13	-25.9	N N	-25.9	5.8	2.019	45	7.2
1000 lb	WME14	- <u>54.7</u>	Y	-3.5	5.8	2.019	45	7.2
1000 lb	WME15	10.0	Ň	10.0	5.8	2.019	45	7.2
1000 lb	WME17	-40.1	Ÿ	5.0	5.8	2.019	45	7.2
1000 lb	WME19	41.9	Ÿ	-1.5	5.8	2.019	45	7.2
1000 lb	WME20	13.3	Ň	13.3	5.8	2.019	45	7.2
1000 lb	WME21	-8.0	Ň	-8.0	5.8	2.019	45	7.2
1000 lb	WME22	-4.1	N	-4.1	5.8	2.019	45	7.2
1000 lb	WME23	-56.7	Ÿ	5.8	5.8	2.019	45	7.2
1000 lb	WME24	-18.8	N N	-18.8	5.8	2.019	45	7.2
1000 ID	VVI'IL∠T	-10.0	1 1	-10.0	5.0	2.013	TJ	/

Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly

Joel P. Lavicky

e-signature is copy only

Metrologist

11/15/2021

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains <u>all</u> pages. This document may not be reproduced except in <u>full</u>, without the written consent of the Nebraska Standards Laboratory.