

DEPARTMENT OF AGRICULTURE

Submitted By:

Nebraska Standards Laboratory

3721 West Cuming St. Lincoln, NE 68524 (402)-471-2087

Director of Agriculture

Sherry Vinton P.O. Box 94947 Lincoln, NE 68509-4947 (402) 471-2341

www.nda.nebraska.gov

2023-134-1

Calibration Certificate of Mass

Calibration Date: October 11, 2023

FSCP Area 10

Lincoln, NE 68524

3721 West Cuming St.

Point of Contact: Andrew Brasch

Certificate Number:

Ph. 402-471-3422

email: andrew.brasch@nebraska.gov

PO Number: N/A

Test Item(s): 47-Avoirdupois weights

ID / Asset Number: Area 10

Material: Cast Iron

Manufacture: Various

Artifact(s) Description:

Date Received: October 6, 2023 Serial Number(s): See Next Page

Class Specification: NIST Class F

Condition: Good (some wear)

Reference Standards Used:

NSL lb standards

Procedure Used:

NIST HB 6969, SOP 8 (2019) Metrologist:

JPL

Equipment Used:

Mettler XP 604 Mettler XPR32003

Environmental Cond.

Temp:

Pressure:

723.64 mmHg

Relative Humidity:

Pertinent Information

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. RED print indicates an out-of-compliance reading. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and the uncertainty exceed 95% of the maximum permissible error. All of the tolerances and design specifications (except density, hardness and magnetism) were evaluated according to ASTM E617 (2018) and/or NIST HB 105-1 (2019) for the artifacts designated class.
 - All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm³ reference mass density and an air density of 1.2 mg/cm3 at 20 °C.
- It is the end user's responsibility to verify that the weights meet the accuracy requirements outlined in NIST Handbook 44 (2022), Appendix A Fundamental Considerations, when using the weights for calibration of commercial (Legal for Trade) scales.

Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (k), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.



DEPARTMENT OF AGRICULTURE

Nebraska Standards Laboratory 3721 West Cuming St. Lincoln, NE 68524

(402)-471-2087

Director of Agriculture Sherry Vinton P.O. Box 94947 Lincoln, NE 68509-4947 (402) 471-2341

www.nda.nebraska.gov

Calibration Date: October 11, 2023				Certificate Number:		: 2023-134-	2023-134-1	
Calibration Results								
Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g) (k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm³)
25 lb	WM-D51	0.98	Υ	-0.66	0.14	2.01	1.1	7.2
25 lb	WM-D52	0.23	N	0.23	0.14	2.01	1.1	7.2
25 lb	WM-D53	0.50	N.	0.50	0.14	2.01	1.1	7.2
25 lb	WM-D54	-0.42	N	-0.42	0.14	2.01	1.1	7.2
25 lb	WM-D55	1.11	Y	0.22	0.14	2.01	1.1	7.2
25 lb 25 lb	WM-D56 WM-D57	0.23 0.82	N N	0.23 0.82	0.14 0.14	2.01 2.01	1.1 1 1	7.2 7.2
25 lb	WM-D58	0.82	N N	0.82 0.41	0.14	2.01	1.1	7.2
25 lb	WM-D59	-0.28	N	-0.28	0.14	2.01	1.1	7.2
25 lb	WM-D60	1.21	Y	-0.12	0.14	2.01	1.1	7.2
25 lb	WMD-61	1.00	Ý	-0.12	0.14	2.01	1.1	7.2
25 lb	D36	-0.26	Ň	-0.26	0.14	2.01	1.1	7.2
50 lb	C-C1	1.86	Ň	1.86	0.28	2.01	2.3	7.2
50 lb	C-C2	-0.06	Ň	-0.06	0.28	2.01	2.3	7.2
50 lb	C-C3	1.19	N	1.19	0.28	2.01	2.3	7.2
50 lb	C-C4	1.81	N	1.81	0.28	2.01	2.3	7.2
50 lb	C-C6	3.81	Y	-0.32	0.28	2.01	2.3	7.2
50 lb	C-C8	4.92	Υ	0.53	0.28	2.01	2.3	7.2
50 lb	C-C10	0.64	N	0.64	0.28	2.01	2.3	7.2
50 lb	C-C13	0.88	N	0.88	0.28	2.01	2.3	7.2
50 lb	<u>C-C14</u>	5.45	Y	0.89	0.28	2.01	2.3	7.2
50 lb	C-C15	2.01	N	2.01	0.28	2.01	2.3	7.2
50 lb	C-C16	0.98	N.	0.98	0.28	2.01	2.3	7.2
50 lb	C-C18	1.08	N N	1.08	0.28	2.01	2.3	7.2 7.2
50 lb	C-C19	0.73 1.47	N N	0.73	0.28 0.28	2.01 2.01	2.3	
50 lb 1000 lb	A5C5 WME1	58.8	Y	1.47 10.9	<u> </u>	2.01	2.3 45	7.2 7.2
1000 lb	WME2	-4.5	N N	-4.5	5.7	2.01	45	7.2
1000 lb	WME3	63.0	Y	3.2	5.7	2.01	45	7.2
1000 lb	WME4	11.3	Ň	11.3	5.7	2.01	45	7.2
1000 lb	WME5	-10.7	Ň	-10.7	5.7	2.01	45	7.2
1000 lb	WME6	16.3	Ň	16.3	5.7	2.01	45	7.2
1000 lb	WME7	-24.8	Ň	-24.8	5.7	2.01	45	7.2
1000 lb	WME9	18.3	N	18.3	5.7	2.01	45	7.2
1000 lb	WME10	-9.8	N	-9.8	5.7	2.01	45	7.2
1000 lb	WME11	-17.9	N	-17.9	5.7	2.01	45	7.2
1000 lb	WME12	-1.5	N	-1.5	5.7	2.01	45	7.2
1000 lb	WME13	3.3	N	3.3	5.7	2.01	45	7.2
1000 lb	WME14	-6.4	N	-6.4	5.7	2.01	45	7.2
1000 lb	WME15	5.9	N	5.9	5.7	2.01	45	7.2
1000 lb	WME17	<u>56.9</u>	<u>Y</u>	4.4	5.7	2.01	4 <u>5</u>	7.2
1000 lb	WME19	- <u>101.2</u>	Y N	5.5	<u>5.7</u>	2.01	45 45	7.2
1000 lb 1000 lb	WME20 WME21	0.3 66.3	N	0.3 18.3	5.7 5.7	2.01 2.01	45 45	7.2 7.2
1000 lb	WME22	-6.3	N N	-6.3	5.7 5.7	2.01	45 45	7.2
1000 lb	WME23	-0.3 50.4	Y	22.3	5.7 5.7	2.01	45 45	7.2
1000 lb	WME24	-0.8	N	-0.8	5.7	2.01	45	7.2

Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly

Joel P. Lavicky Metrologist

10/23/2023 Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in <u>full</u>, without the written consent of the Nebraska Standards Laboratory.