

Office of the Secretary 900 SW Jackson, Room 456 Topeka, Kansas 66612 (785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Expires on: 2/9/2017

Kansas Metrology Laboratory
Calibration Report

Report Number: K15235

Submitted by:

Nebraska Department Of Agriculture Food Safety & Consumer Protection 301 Centennial Mall South Lincoln, NE 68509 Submitted on: 2/8/2016

Vehicle Number: Area 15

Item(s)

	Tested	Adjusted	Rejected
	83	7	2
i			

Quantity	Nominal Mass	Туре
20	25 lb	Weight(s)
2	15 lb	Weight(s)
37	2 lb to 0.001 lb 8 oz to 1/16 oz	Weight Kit
24	2 lb to 0.001 lb 8 oz to 1/16 oz	Weight Kit

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance ±	Expanded Uncertainty (U), (k=2), ±	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
25 lb	WM25-10	11340.55 g	1.1 g	0.17 g	11340.55 g	In Tolerance
25 lb	WM25-100	11340.79 g	1.1 g	0.17 g	11339.90 g	Adjusted
25 lb	WM25-101	11340.34 g	1.1 g	0.17 g	11340.34 g	In Tolerance
25 lb	WM25-102	11340.53 g	1.1 g	0.17 g	11340.53 g	In Tolerance
25 lb	WM25-28	11339.46 g	1.1 g	0.17 g	11339.46 g	In Tolerance
25 lb	WM25-38	11338.58 g	1.1 g	0.17 g	11339.83 g	Adjusted
25 lb	WM25-45	11340.30 g	1.1 g	0.17 g	11340.30 g	In Tolerance
25 lb	WM25-55	11339.27 g	1.1 g	0.17 g	11339.27 g	In Tolerance
25 lb	WM25-56	11339.58 g	1.1 g	0.17 g	11339.58 g	In Tolerance
25 lb	WM25-57	11338.76 g	1.1 g	0.17 g	11339.92 g	Adjusted
25 lb	WM25-58	11338.85 g	1.1 g	0.17 g	11340.00 g	Adjusted
25 lb	WM25-59	11338.84 g	1.1 g	0.17 g	11339.99 g	Adjusted
25 lb	WM25-75	11339.12 g	1.1 g	0.17 g	11339.12 g	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb 28.349523125 g = 1 oz Expires on: 2/9/2017

Kansas Metrology Laboratory

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance ±	Expanded Uncertainty (U), (k=2), ±	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
25 lb	WM25-76	11338.80 g	1.1 g	0.17 g	11340.05 g	Adjusted
25 lb	WM25-77	11339.54 g	1.1 g	0.17 g	11339.54 g	In Tolerance
25 lb	WM25-78	11339.73 g	1.1 g	0.17 g	11339.73 g	In Tolerance
25 lb	WM25-79	11338.65 g	1.1 g	0.17 g	11339.71 g	Adjusted
25 lb	WM25-96	11340.48 g	1.1 g	0.17 g	11340.48 g	In Tolerance
25 lb	WM25-97	11340.42 g	1.1 g	0.17 g	11340.42 g	In Tolerance
25 lb	WM25-98	11340.54 g	1.1 g	0.17 g	11340.54 g	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb

28.349523125 g = 1 oz

Report Number: K15235

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance ±	Expanded Uncertainty (U), (k=2), ±	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
15 lb	WM15-3	6803.922 g	0.68 g	0.090 g	6803.922 g	In Tolerance
15 lb	WM15-4	6803.432 g	0.68 g	0.090 g	6803.432 g	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb

28.349523125 g = 1 oz

Kansas Metrology Laboratory

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance ±	Expanded Uncertainty (U), (k=2), ±	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
2 lb	3A11 1	907.119 g	0.091 g	0.011 g	907.119 g	In Tolerance
2 lb	3A11 2	907.131 g	0.091 g	0.011 g	907.131 g	In Tolerance
2 lb	3A113	907.144 g	0.091 g	0.011 g	907.144 g	In Tolerance
2 lb	3A11 4	907.120 g	0.091 g	0.011 g	907.120 g	In Tolerance
2 lb	3A115	907.137 g	0.091 g	0.011 g	907.137 g	In Tolerance
2 lb	3A11 6	907.123 g	0.091 g	0.011 g	907.123 g	In Toleranc
2 lb	3A11 7	907.132 g	0.091 g	0.011 g	907.132 g	In Tolerance
2 lb	3A11 8	907.132 g	0.091 g	0.011 g	907.132 g	In Tolerance
2 lb	3A11 9	907.129 g	0.091 g	0.011 g	907.129 g	In Tolerance
2 lb	3A11 10	907.143 g	0.091 g	0.011 g	907.143 g	In Toleranc
2 lb	3A11 11	907.129 g	0.091 g	0.011 g	907.129 g	In Tolerance
2 lb	3A11 12	907.123 g	0.091 g	0.011 g	907.123 g	In Tolerance
2 lb	3A11 13	907.134 g	0.091 g	0.011 g	907.134 g	In Tolerance
2 lb	3A11 14	907.148 g	0.091 g	0.011 g	907.148 g	In Tolerance
1 lb	3A11 15	453.5749 g	0.070 g	0.0084 g	453.5749 g	In Tolerance
1 lb	3A11 16	453.5509 g	0.070 g	0.0084 g	453.5509 g	In Tolerance
0.3 lb	3A11	136.0890 g	0.027 g	0.0032 g	136.0890 g	In Tolerance
0.2 lb	3A11	90.7135 g	0.018 g	0.0021 g	90.7135 g	In Tolerance
0.1 lb	3A11	45.3527 g	0.0091 g	0.0011 g	45.3527 g	In Tolerance
0.05 lb	3A11	22.67959 g	0.0045 g	0.00055 g	22.67959 g	In Tolerance
0.03 lb	3A11	13.60663 g	0.0027 g	0.00032 g	13.60663 g	In Tolerance
0.02 lb	3A11	9.07276 g	0.0018 g	0.00022 g	9.07276 g	In Tolerance
0.01 lb	3A11	4.53723 g	0.0015 g	0.00018 g	4.53723 g	In Tolerance
0.005 lb	3A11	2.26826 g	0.0012 g	0.00015 g	2.26826 g	In Tolerance
0.003 lb	3A11	1.36033 g	0.00099 g	0.00012 g	1.36033 g	In Tolerance
0.002 lb	3A11	0.90772 g	0.00087 g	0.00011 g	0.90772 g	In Tolerance
0.001 lb	3A11	0.453022 g	0.00070 g	0.000094 g	0.453022 g	In Tolerance
0.001 lb	3A11 •	0.453882 g	0.00070 g	0.000094 g	0.453882 g	In Tolerance
8 oz	3A11	226.7845 g	0.045 g	0.0053 g	226.7845 g	In Tolerance
4 oz	3A11	113.3994 g	0.023 g	0.0028 g	113.3994 g	In Tolerance
2 oz	3A11	56.6953 g	0.011 g	0.0013 g	56.6953 g	In Tolerance
1 oz	3A11	28.34446 g	0.0054 g	0.00065 g	28.34446 g	Rejected
1/2 oz	3A11	14.17316 g	0.0028 g	0.00033 g	14.17316 g	In Tolerance
1/4 oz	3A11	7.08675 g	0.0017 g	0.00020 g	7.08675 g	In Tolerance
1/8 oz	3A11	3.54244 g	0.0013 g	0.00016 g	3.54244 g	Rejected
1/16 oz	3A11	1.77247 g	0.0011 g	0.00014 g	1.77247 g	In Tolerance
1/16 oz	3A11 •	1.77201 g	0.0011 g	0.00014 g	1.77201 g	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb 28.349523125 g = 1 oz

Report Number: K15235

Expires on: 2/9/2017

Kansas Metrology Laboratory

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance ±	Expanded Uncertainty (U), (k=2), ±	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
2 lb	9-OPI-5 1	907.181 g	0.091 g	0.011 g	907.181 g	In Tolerance
2 lb	9-OPI-5 2	907.176 g	0.091 g	0.011 g	907.176 g	In Tolerance
2 lb	9-OPI-5 3	907.172 g	0.091 g	0.011 g	907.172 g	In Tolerance
1 lb	9-OPI-5	453.5879 g	0.070 g	0.0084 g	453.5879 g	In Tolerance
8 oz	9-OPI-5	226.7925 g	0.045 g	0.0053 g	226.7925 g	In Tolerance
4 oz	9-OPI-5	113.3969 g	0.023 g	0.0028 g	113.3969 g	In Tolerance
2 oz	9-OPI-5	56.6981 g	0.011 g	0.0013 g	56.6981 g	In Tolerance
1 oz	9-OPI-5	28.34982 g	0.0054 g	0.00065 g	28.34982 g	In Tolerance
1/2 oz	9-OPI-5	14.17612 g	0.0028 g	0.00033 g	14.17612 g	In Tolerance
1/4 oz	9-OPI-5	7.08756 g	0.0017 g	0.00020 g	7.08756 g	In Tolerance
1/8 oz	9-OPI-5	3.54286 g	0.0013 g	0.00016 g	3.54286 g	In Tolerance
1/16 oz	9-OPI-5	1.77164 g	0.0011 g	0.00014 g	1.77164 g	In Tolerance
1/16 oz	9-OPI-5 •	1.77183 g	0.0011 g	0.00014 g	1.77183 g	In Tolerance
0.2 lb	N-99-C	90.7249 g	0.018 g	0.0021 g	90.7249 g	In Tolerance
0.2 lb	N-99-C •	90.7264 g	0.018 g	0.0021 g	90.7264 g	In Tolerance
0.1 lb	N-99-C	45.3629 g	0.0091 g	0.00 1 1 g	45.3629 g	In Tolerance
0.05 lb	N-99-C	22.68075 g	0.0045 g	0.00055 g	22.68075 g	In Tolerance
0.02 lb	N-99-C	9.07114 g	0.0018 g	0.00022 g	9.07114 g	In Tolerance
0.02 lb	N-99-C •	9.07172 g	0.0018 g	0.00022 g	9.07172 g	In Tolerance
0.01 lb	N-99-C	4.53576 g	0.0015 g	0.00018 g	4.53576 g	In Tolerance
0.005 lb	N-99-C	2.26844 g	0.0012 g	0.00015 g	2.26844 g	In Tolerance
0.002 lb	N-99-C	0.90731 g	0.00087 g	0.00011 g	0.90731 g	In Tolerance
0.002 lb	N-99-C •	0.90730 g	0.00087 g	0.00011 g	0.90730 g	In Tolerance
0.001 lb	N-99-C	0.453662 g	0.00070 g	0.000094 g	0.453662 g	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb

28.349523125 g = 1 oz

Report Number: K15235

Expires on: 2/9/2017

Kansas Metrology Laboratory

Uncertainty Statement:

The combined standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, the standard uncertainty for any uncorrected errors associated with buoyancy corrections(applies to mass values only), the standard uncertainty for any uncorrected errors associated with temperature correction(applies to length and volume values only), and a component of uncertainty to account for any observed deviations from NIST(The National Institute of Standards and Technology) values that are less than surveillance limits. The combined standard uncertainty is multiplied by a coverage factor of 2 to give an expanded uncertainty, which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 1993 ISO Guide to the Expression of Uncertainty in Measurement and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

Traceability Statement:

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

Condition of Item(s) Submitted for Testing:

Minor wear.

Treatment of Item(s) before Testing:

Item(s) were tested as found.

Documentary Standards:

NIST Handbook 105 Series

NISTIR 6969: SOP 8, SOP 4, and/or SOP 7

ASTM E 617-13 or OIML R 111-1 2004(E)

Environmental Conditions:

Temperature:

20.0 °C

Barometric Pressure:

730.08 mmHg

Relative Humidity:

44.4 %

Test Date: 2/10/2016

Due Date: 2/9/2017 -Per state statute K.S.A. 83-304(a).

Keith Arkenberg, Metrologist

2/15/2016



This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.

Report Number: K15235



Office of the Secretary 900 SW Jackson, Room 456 Topeka, Kansas 66612 (785) 296-3556

Governor Sam Brownback

Report Number: K15235-TM

Submitted on: 2/8/2016

Jackie McClaskey, Secretary

Kansas Metrology Laboratory
Calibration Report

Submitted by:

Expires on: 2/9/2017

Nebraska Department Of Agriculture Food Safety & Consumer Protection P.O. Box 94757 Lincoln, NE 68509

Reference Number: AREA 15

Item(s)

Tested	Adjusted	Rejected
5	2	0

Quantity	Nominal Volume	Type
3	5 gal	Bottom Drop Test Measure "To Deliver"
2	5 gal	Handheld Test Measure "To Deliver"

The calibration of items is performed according to NISTIR 7383, SOP 19 Volume Transfer. Tolerances are applied from NISTHB 105-3. The volume applies when a 10 second drain is observed for 5 gallon hand held test measures. For 5 gallon bottom drop test measures and provers a 30 second drain applies. The drain time starts when the cessation of the main flow is observed.

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (/°F)	Volume as Found @ 60 °F	Tolerance ±	Expanded Uncertainty (U), (k=2.02), ±	Volume as Left @ 60 °F	Adjusted/ In Tolerance/ Rejected
5 gal	05-40547-01	Stainless Steel	0.0000265	5.00006 gal	0.00250 gal	0.00084 gal	5.00006 gal	In Tolerance
5 gal	05-40547-02	Stainless Steel	0.0000265	4.99971 gal	0.00250 gal	0.00084 gal	4.99971 gal	In Tolerance
5 gal	05-40547-03	Stainless Steel	0.0000265	4.99971 gal	0.00250 gal	0.00084 gal	4.99971 gal	In Tolerance
5 gal	39423 A	Stainless Steel	0.0000265	5.00274 gal	0.00250 gal	0.00085 gal	4.99993 gal	Adjusted
5 gal	39423 D	Stainless Steel	0.0000265	5.00490 gal	0.00250 gal	0.00085 gal	5.00058 gal	Adjusted
:								

The data in the above table of this report only applies to those items specifically listed on this report.

1 m³=1 000 L=264.1720 gal

Temperature Corrections

ltem

Item	Temperature °F	in³
	-20	-2.45
	-15	-2.30
d)	-10	-2.14
ure	-5	-1.99
sas	0	-1.84
Ĕ	5	-1.68
st	10	-1.53
٦e	15	-1.38
<u>e</u>	20	-1.22
Temperature Correction for 5 gal Stainless Steel Test Measure (CCE= 0.0000265/°F)	25	-1.07
:55	30	-0.92
on for 5 gal Stainles CCE= 0.0000265/°F)	35	-0.77
tai 265	40	-0.61
1 S	45	-0.46
.8 .900	50	-0.31
r 5 0.0	55	-0.15
- <u>1</u> 1-	60	0.00
io C	65	0.15
ecti (70	0.31
rre	75	0.46
පි	80	0.61
<u>=</u>	85	0.77
atn	90	0.92
era	95	1.07
ω	100	1.22
<u> </u>	105	1.38
•	110	1.53
	115	1.68
	120	1.84

	-15	-1.61
- e	-10	-1.50
nsı	-5	-1.40
le3	0	-1.29
<u> </u>	5	-1.18
es.	10	-1.07
	15	-0.97
tee	20	-0.86
l S	25	-0.75
000	30	-0.64
on for 5 gal Low Carb (CCE= 0.0000186/°F)	35	-0.54
√ C 86,	40	-0.43
010	45	-0.32
a 000	50	-0.21
5.8	55	-0.11
占出	60	0.00
	65	0.11
Liol (70	0.21
ect	75	0.32
ž	80	0.43
٥	85	0.54
ure	90	0.64
at.	95	0.75
Temperature Correction for 5 gal Low Carbon Steel Test Measure (CCE= 0.0000186/°F)	100	0.86
<u>E</u>	105	0.97
Te	110	1.07
	115	1.18
	120	1.29

Temperature °F -20

-1.72

CCE = Coefficient of Cubical Expansion

Kansas Metrology Laboratory Report Number: K15235-TM Expires on: 2/9/2017

Uncertainty Statement:

The combined standard uncertainty includes the standard uncertainty reported for the standards, the standard uncertainty for the measurement process, the standard uncertainty for the water density equation (Metrologia Tanaka, et al), the standard uncertainty for any uncorrected errors associated with temperature correction (applies to length and volume values only), the standard uncertainty for reading the meniscus (when applicable), the standard uncertainty for viscosity, and a component of uncertainty to account for any observed deviations from NIST(The National Institute of Standards and Technology) values that are less than surveillance limits. The combined standard uncertainty is multiplied by the coverage factor (k-value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The k-value reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is consistent with the 1993 ISO Guide to the Expression of Uncertainty in Measurement and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

Traceability Statement:

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

Condition of Item(s) Submitted for Testing:

Minor wear.

Treatment of Item(s) before Testing:

Item(s) were tested as found.

Water Temperature at Time of Test:

70.50 °F

Documentary Standards:

-NIST Handbook 105-3 (2010) -NISTIR 7383 (2013), SOP 19

Environmental Conditions:

21.56 °C Temperature: 730.41 mmHg Barometric Pressure:

Relative Humidity: 23.1 %

Test Date: 2/10/2016

2/9/2017 Due Date: -Per state statute K.S.A. 83-304(a).

2/15/2016 Keith Arkenberg, Metrologist

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.



Office of the Secretary 900 SW Jackson, Room 456 Topeka, Kansas 66612 (785) 296-3556

Jackie McClaskey, Secretary

Test Date: 2/11/2016

Kansas Metrology Laboratory
Certificate of Calibration

Governor Sam Brownback Test No.: K15235-1.0



Nebraska Department Of Agriculture Food Safety & Consumer Protection 301 Centennial Mall South Lincoln NE 68509

> Manufacturer: Rice Lake S/N: K5 and K6 Number of Pieces: 2

Nominal Mass	Weight's Markings	Assumed Density (g/cm³)	Conventional Mass As Found (g)	Conventional Mass As Left (g)	Expanded Uncertainty ± (mg)	In Tolerance Adjusted Rejected
2 kg	2 kg K5	7.95	2000.0004	2000.0004	3.7	In Tolerance
2 kg	2 kg K6	7.95	2000.0061	2000.0061	3.7	In Tolerance
4 kg	4 kg Combination*	7.95	4000.0051	4000.0051	3.7	In Tolerance

This document certifies the above mentioned artifacts were compared to the Standards of the State of Kansas which are traceable to the National Institute of Standards and Technology. The conventional mass is the weight in normal air (1.2 mg/cm³) at 20 °C versus the reference density of 8.0 g/cm³ Calibration of listed items was performed according to NISTIR 6969, SOP 4 (Double Substitution) and/or NISTIR 5672, SOP 5 (3-1).

Tolerances were evaluated to ASTM Class 4. Surface finish and magnetism were not evaluted as it is assumed to be done by the manufacturer.

^{*} The 4 kg combination is a grouping of the two 2 kg standards listed above, K5 and K6

Test Date: 2/11/2016

Kansas Metrology Laboratory

Uncertainty Statement:

The combined standard uncertainty includes the standard uncertainty reported for the; standards, tare weights, the standard uncertainty for the measurement process, the standard uncertainty for air buoyancy corrections as stated in OIML R111-1 [2004E] eq. C.6.3-1, and a component of uncertainty to account for any observed deviations (Bias) from NIST (National Institute of Standards and Technology) values that are less than surveillance limits. Factors not considered in the evaluation: magnetism, weights are considered to meet magnetism specifications unless measurement aberrations are noted, balance eccentricity and linearity, these factors are considered as a part of the measurement assurance process when using a check standard with adequate degrees of freedom. The combined uncertainty is multiplied by the coverage factor (k-value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The coverage factor reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is also consistent with and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

Uncertainty Analysis:

Nominal	S_p	U _{S (k =1)}	$U_{tare(k=1)}$	U _{Air Buoyancy Eq.}	ρ_{air}	Procedure
2 kg	1.68	0.252	No Tare	-0.0297	1.16795	SOP 4
2 kg	1.68	0.252	No Tare	~0.0298	1.16805	SOP 4
4 kg	1.68	0.310	0.0535	-0.0741	1.16741	SOP 4
			······			
			•••••			
					* * * *	
	All values list	ed as a compone	nt of the overall u	incertainty are in uni	ts of milligrams	(mg) or (mg/cm³).

Traceability Statement:

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

Condition of Item(s) Submitted for Testing: Minor wear.

Treatment of Item(s) before Testing: Item(s) were tested as found.

Documentary Standards: NIST Handbook 105 Series, NISTIR 6969, SOP 4, NISTIR 5672, SOP 5, & ASTM E 617-13 or OIML R111-1

Item(s) Received on: 2/8/2016

Item(s) Acclimated: 2/8/2016 4:25:00 PM

Environmental Conditions: Temperature Barometric Pressure Relative Humidity
20.3 °C 740.96 mmHg 45.4 %

Values are averages recorded over the duration of testing

2/15/2016

Keith Arkenberg, Metrologist

Date

KML Software Version: 8.3

Test No.: K15235-1.0

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.



Office of the Secretary 900 SW Jackson, Room 456 Topeka, Kansas 66612 (785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Test Date: 2/12/2016

Kansas Metrology Laboratory Certificate of Calibration

Test No.: K15235-1.1



Nebraska Department Of Agriculture Food Safety & Consumer Protection 301 Centennial Mall South Lincoln, NE 68509

Manufacturer: Troemner S/N: FSCP-3A14

Number of Pieces: 13 of 25 Total

Nominal Mass	Weight's Markings	Assumed Density (g/cm³)	Conventional Mass As Found (g)	Conventional Mass As Left (g)	Expanded Uncertainty ± (mg)	In Tolerance Adjusted Rejected
1 kg	1	8.03	1000.00185	1000.00185	0.16	In Tolerance
500 g	500	8.03	500.00088	500.00088	0.13	In Tolerance
200 g	200	8.03	200.000327	200.000327	0.093	In Tolerance
200 g	200 •	8.03	200.000326	200.000326	0.093	In Tolerance
100 g	100	8.03	100.000141	100.000141	0.018	In Tolerance
50 g	50	8.03	49.999978	49.999978	0.010	In Tolerance
20 g	20	8.03	20.0001002	20.0001002	0.0055	In Tolerance
20 g	20 •	8.03	20.0000213	20.0000213	0.0055	In Tolerance
10 g	10	8.03	10.0000453	10.0000453	0.0050	In Tolerance
5 g	5	8.03	5.0000288	5.0000288	0.0024	In Tolerance
2 g	2	8.03	2.0000256	2.0000256	0.0024	In Tolerance
2 g	2 •	8.03	2.0000129	2.0000129	0.0024	In Tolerance
1 g	1	8.03	1.0000310	1.0000310	0.0013	In Tolerance

This document certifies the above mentioned artifacts were compared to the Standards of the State of Kansas which are traceable to the National Institute of Standards and Technology. The conventional mass is the weight in normal air (1.2 mg/cm³) at 20 °C versus the reference density of 8.0 g/cm³ Calibration of listed items was performed according to NISTIR 6969, SOP 4 (Double Substitution) and/or NISTIR 5672, SOP 5 (3-1).

Tolerances were evaluated to OIML Class F₁. Surface finish and magnetism were not evaluted as it is assumed to be done by the manufacturer.

Test Date: 2/12/2016

Kansas Metrology Laboratory

Test No.: K15235-1.1

Uncertainty Statement:

The combined standard uncertainty includes the standard uncertainty reported for the; standards, tare weights, the standard uncertainty for the measurement process, the standard uncertainty for air buoyancy corrections as stated in OIML R111-1 [2004E] eq. C.6.3-1, and a component of uncertainty to account for any observed deviations (Bias) from NIST (National Institute of Standards and Technology) values that are less than surveillance limits. Factors not considered in the evaluation: magnetism, weights are considered to meet magnetism specifications unless measurement aberrations are noted, balance eccentricity and linearity, these factors are considered as a part of the measurement assurance process when using a check standard with adequate degrees of freedom. The combined uncertainty is multiplied by the coverage factor (k-value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The coverage factor reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is also consistent with and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

Uncertainty Analysis:

Nominal	S_p	$U_{S(k=1)}$	U _{tare (k =1)}	U _{Air Buoyancy Eq.}	$ ho_{air}$	Procedure
1 kg	0.0553	0.0535	No Tare	-0.00997	1.17501	SOP 5
500 g	0.0538	0.0273	No Tare	-0.00776	1.17577	SOP 5
200 g	0.0408	0.0121	No Tare	-0.00312	1.17503	SOP 5
200 g	0.0408	0.0121	No Tare	-0.00311	1.17564	SOP 5
100 g	0.00151	0.00907	No Tare	-0.00156	1.17530	SOP 5
50 g	0.00238	0.00470	No Tare	-0.000750	1.17593	SOP 5
20 g	0.00158	0.00222	No Tare	-0.000298	1.17689	SOP 5
20 g	0.00158	0.00222	No Tare	-0.000299	1.17649	SOP 5
10 g	0.00160	0.00183	No Tare	-0.000151	1.17505	SOP 5
5 g	0.000695	0.000970	No Tare	~0.0000742	1.17449	SOP 5
2 g	0.00102	0.000495	No Tare	-0.0000294	1.17661	SOP 5
2 g	0.00102	0.000495	No Tare	-0.0000293	1.17722	SOP 5
1 g	0.000459	0.000455	No Tare	-0.0000148	1.17597	SOP 5

All values listed as a component of the overall uncertainty are in units of milligrams (mg) or (mg/cm3).

Traceability Statement:

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

Condition of Item(s) Submitted for Testing:

Minor wear.

Treatment of item(s) before Testing:

Weights had visible finger prints that were wiped clean using a dry cloth

Documentary Standards:

NIST Handbook 105 Series, NISTIR 6969, SOP 4, NISTIR 5672, SOP 5, & ASTM E 617-13 or OIML R111-1

Item(s) Received on:

2/8/2016

Item(s) Acclimated:

2/8/2016 4:25:00 PM

Environmental Conditions:

Temperature	Barometric Pressure	Relative Humidity
20.1 °C	744.61 mmHg	46.9 %

Values are averages recorded over the duration of testing

Keith Arkenberg , Metrologist

2/15/2016 Date

KML Software Version: 9.2

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.



Office of the Secretary 900 SW Jackson, Room 456 Topeka, Kansas 66612 (785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Test Date: 2/15/2016

Kansas Metrology Laboratory Certificate of Calibration

Test No.: K15235-1.2



Nebraska Department Of Agriculture Food Safety & Consumer Protection 301 Centennial Mall South Lincoln NE 68509

> Manufacturer: Troemner S/N: FSCP-3A14 Number of Pieces: 12 of 25 total

Nominal Mass	Weight's Markings	Assumed Density (g/cm ³)	Conventional Mass As Found (g)	Conventional Mass As Left (g)	Expanded Uncertainty ± (mg)	In Tolerance Adjusted Rejected
500 mg	No Markings	7.95	0.5000137	0.5000137	0.0012	In Tolerance
200 mg	No Markings	7.95	0.20001474	0.20001474	0.00066	In Tolerance
200 mg	•	7.95	0.20001305	0.20001305	0.00066	In Tolerance
100 mg	No Markings	7.95	0.10001190	0.10001190	0.00073	In Tolerance
50 mg	No Markings	7.95	0.04999870	0.04999870	0.00042	In Tolerance
20 mg	No Markings	7.95	0.02001698	0.02001698	0.00028	In Tolerance
20 mg	•	7.95	0.02000207	0.02000207	0.00028	In Tolerance
10 mg	No Markings	7.95	0.01000801	0.01000801	0.00039	In Tolerance
5 mg	No Markings	2.7	0.00499804	0.00499804	0.00034	In Tolerance
2 mg	No Markings	2.7	0.00201004	0.00201004	0.00031	In Tolerance
2 mg	•	2.7	0.00200547	0.00200547	0.00031	In Tolerance
1 mg	No Markings	2.7	0.00101183	0.00101183	0.00039	In Tolerance

This document certifies the above mentioned artifacts were compared to the Standards of the State of Kansas which are traceable to the National Institute of Standards and Technology. The conventional mass is the weight in normal air (1.2 mg/cm³) at 20 °C versus the reference density of 8.0 g/cm³ Calibration of listed items was performed according to NISTIR 6969, SOP 4 (Double Substitution) and/or NISTIR 5672, SOP 5 (3-1).

Tolerances were evaluated to OIML Class F1. Surface finish and magnetism were not evaluted as it is assumed to be done by the manufacturer.

Test Date: 2/15/2016

Kansas Metrology Laboratory

Uncertainty Statement:

The combined standard uncertainty includes the standard uncertainty reported for the; standards, tare weights, the standard uncertainty for the measurement process, the standard uncertainty for air buoyancy corrections as stated in OIML R111-1 [2004E] eq. C.6.3-1, and a component of uncertainty to account for any observed deviations (Bias) from NIST (National Institute of Standards and Technology) values that are less than surveillance limits. Factors not considered in the evaluation: magnetism, weights are considered to meet magnetism specifications unless measurement aberrations are noted, balance eccentricity and linearity, these factors are considered as a part of the measurement assurance process when using a check standard with adequate degrees of freedom. The combined uncertainty is multiplied by the coverage factor (k-value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The coverage factor reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is also consistent with and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

Uncertainty Analysis:

Nominal	S_p	$u_{S(k=1)}$	U _{tare (k =1)}	UAir Buoyancy Eq.	$ ho_{air}$	Procedure
500 mg	0.000503	0.000255	No Tare	0.0000172	1.15071	SOP 5
200 mg	0.000284	0.000150	No Tare	0.00000685	1.15210	SOP 5
200 mg	0.000284	0.000150	No Tare	0.00000683	1.15250	SOP 5
100 mg	0.000324	0.000150	No Tare	0.00000343	1.15206	SOP 5
50 mg	0.000180	0.0000950	No Tare	0.00000172	1.15098	SOP 5
20 mg	0.000117	0.0000650	No Tare	0.000000686	1.15196	SOP 5
20 mg	0.000117	0.0000650	No Tare	0.000000684	1.15237	SOP 5
10 mg	0.000179	0.0000700	No Tare	0.000000346	1.15136	SOP 5
5 mg	0.000149	0.0000550	No Tare	0.00000546	1.15037	SOP 5
2 mg	0.000140	0.0000550	No Tare	0.00000219	1.15108	SOP 5
2 mg	0.000140	0.0000550	No Tare	0.00000219	1.15141	SOP 5
1 mg	0.000177	0.0000650	No Tare	0.00000109	1.15165	SOP 5

All values listed as a component of the overall uncertainty are in units of milligrams (mg) or (mg/cm³).

Traceability Statement:

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

Condition of Item(s) Submitted for Testing: Minor wear.

Treatment of Item(s) before Testing: Item(s) were tested as found.

NIST Handbook 105 Series, NISTIR 6969, SOP 4, NISTIR 5672, SOP 5, & ASTM E 617-13 or OIML R111-1 **Documentary Standards:**

Item(s) Received on: 2/8/2016

Item(s) Acclimated: 2/8/2016 4:25:00 PM

Environmental Conditions: Relative Humidity Barometric Pressure Temperature 20.0 °C 729.82 mmHg 42.5 %

Values are averages recorded over the duration of testing

2/15/2016 Keith Arkenberg, Metrologist

KML Software Version: 8.3

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.

Date

Test No.: K15235-1.2

raceable Salibration CERTIFICATE RICE LAKE

Traceable Report Number:

2403877

STATE OF NEBRASKA

PO BOX 94757

LINCOLN, NE 68509-0757

Purchase Order Number:

4024165256

Client:

Contractor:

STATE OF NEBRASKA

PO BOX 94757

LINCOLN, NE 68509-0757

Date Received:

22 Feb 2016

Date Calibrated:

26 Feb 2016

Temperature Range:

Contractor Requested Recall Date:

1 year 19.98 °C

Pressure Range:

735.06 mmHg

Relative Humidity Range:

57.20 %

Air Density Range:

1.1591 mg/cm3

NIST Certificate Number:

684/286541-15,684/284451-14

Although there are two NIST numbers, one or both may apply.

Tested By:

Procedure:

Modified Substitution (WI05-0023)

Description of Weights:

1/8 oz Satin Finish Weight, NIST Class F, S/N 6GTR.

Conventional Mass Corr.

Nominal Value	ID	As Found (mg)	As Found In Tol	As Left (mg)	As Left In Tol	Unc. (mg)	k	Tol.* (mg)	Balance Used	Standard Set Used	Assumed Density (g/cm ₃)
0.125 oz	6GTR.	0.26	Υ	0.26	Υ	0.16	2	1.3	638Q	D564Q	7.84

E WEIGHING

ANSI/NCSL Z540-1-1994; Part 1 & ISO/IEC 17025 Accredited

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Check with your local state agency for certification of compliance on Legal for Trade items. The weight tolerance class is referenced in the Description of Weights. Unless otherwise noted, weights tested meet the requirements of the class. *The specifications for the weight classes can be found in NIST Handbook 105-1, ASTM E-617 or OIML R111.

Prepared By:

Rice Lake Weighing Systems

230 West Coleman Street, Rice Lake, WI 54868 • USA TEL: 715-234-9171 • FAX: 715-234-6967 • www.ricelake.com An ISO 9001 registered company

Page 1 of 1

26 Feb 2016

Dan Demers, Metrologist



RICELAKE Fraceable Galibration CERTIFICATE

Traceable Report Number:

2403877A

Contractor: STATE OF NEBRASKA

PO BOX 94757

LINCOLN, NE 68509-0757

Purchase Order Number:

4024165256

Client:

STATE OF NEBRASKA

PO BOX 94757

LINCOLN, NE 68509-0757

Date Received:

22 Feb 2016

Date Calibrated:
Contractor Requested Recall Date:

25 Feb 2016

Temperature Range:

1 year

20.12 °C

Pressure Range:

735.17 mmHg

Relative Humidity Range:

57.84 %

Air Density Range:

1.1586 mg/cm3

NIST Certificate Number:

684/286541-15,684/284451-14

Although there are two NIST numbers, one or both may apply.

Tested By:

22

Procedure:

Modified Substitution (WI05-0023)

Description of Weights:

1 oz Satin Finish Weight, NIST Class F, S/N 6GTP.

TMAP

TMAP

TMAP

TMAP

TMAP

ANSI/NCSL Z540-1-1994; Part 1 & ISO/IEC 17025 Accredited

Conventional Mass Corr.

		Conventional Mass Con.										
Nominal Value	ID	As Found (mg)	As Found In Tol	As Left (mg)	As Left In Tol		Unc. (mg)	k	Tol.* (mg)	Balance Used	Standard Set Used	Assumed Density (g/cm ₃)
1 oz	6GTP.	1.43	Υ	1.43	Υ		0.67	2	5.4	1221Q	D564Q	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Check with your local state agency for certification of compliance on Legal for Trade items.

The weight tolerance class is referenced in the Description of Weights. Unless otherwise noted, weights tested meet the requirements of the class.

*The specifications for the weight classes can be found in NIST Handbook 105-1, ASTM E-617 or OIML R111.

Prepared By:

Rice Lake Weighing Systems

230 West Coleman Street, Rice Lake, WI 54868 • USA TEL: 715-234-9171 • FAX: 715-234-6967 • **www.ricelake.com** An ISO 9001 registered company

Page 1 of 1

Dated 26 Feb 2016

Dan Demers, Metrologist

