

### Calibration Certificate of Mass

Calibration Date: July 18, 2018

Certificate Number: 2018-063-1

Submitted By: FSCP Area 20  
944 N 20th Rd  
Unadilla, NE 68454

Point of Contact: Kurt Wenninghoff  
Ph. 402-471-3422  
email: kurt.wenninghoff@nebraska.gov  
PO Number:

Test Item(s): (20)-25, (10)-50 & (20)-1000 lb weights

Date Received: July 13, 2018

Serial Number(s): See Next Page

Artifact(s) Description:

ID / Asset Number: N/A

Manufacture: Various

Class Specification: NIST Class F

Condition: Fair (significant wear)

Material: Cast iron

Reference Standards Used:

Procedure Used:

Equipment Used:

NSL lb standards

NIST HB 6969, SOP 8

Mettler KA30-3

Metrologist:

JPL

Mettler XP 604

Environmental Cond.

Temp: 25.5 °C

Pressure: 764.032 mmHg

Relative Humidity:

48.8 %

Pertinent Information

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. RED print indicates an out-of-compliance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (k), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: July 18, 2018

Certificate Number: 2018-063-1


### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
25 lb	WM25-1	-1.40	y	0.41	0.14	2	1.1	7.2
25 lb	WM25-1S	-1.78	y	0.14	0.14	2	1.1	7.2
25 lb	WM25-2	-0.33	y	0.08	0.14	2	1.1	7.2
25 lb	WM25-3	0.08	y	-0.16	0.14	2	1.1	7.2
25 lb	WM25-4	-0.16	y	0.01	0.14	2	1.1	7.2
25 lb	WM25-5	0.01	y	0.58	0.14	2	1.1	7.2
25 lb	WM25-6	0.58	y	0.26	0.14	2	1.1	7.2
25 lb	WM25-7	0.26	y	-0.63	0.14	2	1.1	7.2
25 lb	WM25-8	-0.63	y	-0.13	0.14	2	1.1	7.2
25 lb	WM25-10	-0.13	y	-0.13	0.14	2	1.1	7.2
25 lb	WM25-11	-0.86	n	-0.86	0.14	2	1.1	7.2
25 lb	WM25-12	0.02	n	0.02	0.14	2	1.1	7.2
25 lb	WM25-13	-0.53	n	-0.53	0.14	2	1.1	7.2
25 lb	WM25-14	-0.44	n	-0.44	0.14	2	1.1	7.2
25 lb	WM25-15	-0.57	n	-0.57	0.14	2	1.1	7.2
25 lb	WM25-16	-1.14	y	-0.18	0.14	2	1.1	7.2
25 lb	WM25-17	-0.71	n	-0.71	0.14	2	1.1	7.2
25 lb	WM25-18	-0.23	n	-0.23	0.14	2	1.1	7.2
25 lb	WM25-19	-0.66	n	-0.66	0.14	2	1.1	7.2
25 lb	WM25-20	-0.99	y	0.06	0.14	2	1.1	7.2
50 lb	A5C*13	1.75	n	1.75	0.28	2	2.3	7.2
50 lb	A5C*20	1.02	n	1.02	0.28	2	2.3	7.2
50 lb	WM-C-A1	1.73	n	1.73	0.28	2	2.3	7.2
50 lb	WM-C-A2	-5.90	y	-0.33	0.28	2	2.3	7.2
50 lb	WM-C-A3	-7.35	y	-0.18	0.28	2	2.3	7.2
50 lb	WM-C-A5	2.46	y	1.10	0.28	2	2.3	7.2
50 lb	WM-C-A9	-7.32	y	-0.45	0.28	2	2.3	7.2
50 lb	WM-C-A10	-7.32	y	-0.28	0.28	2	2.3	7.2
50 lb	WM-OPI-C81	1.67	n	1.67	0.28	2	2.3	7.2
50 lb	WM-OPI-C85	2.72	y	1.11	0.28	2	2.3	7.2
1000 lb	2189	-14.4	n	-14.4	5.8	2	45	7.2
1000 lb	2190	-10.8	n	-10.8	5.8	2	45	7.2
1000 lb	2191	-39.6	y	7.8	5.8	2	45	7.2
1000 lb	2192	-7.1	n	-7.1	5.8	2	45	7.2
1000 lb	2194	-14.0	n	-14.0	5.8	2	45	7.2
1000 lb	2195	3.5	n	3.5	5.8	2	45	7.2
1000 lb	2196	-16.0	n	-16.0	5.8	2	45	7.2
1000 lb	2197	-25.0	n	-25.0	5.8	2	45	7.2
1000 lb	2198	-5.2	n	-5.2	5.8	2	45	7.2
1000 lb	A-1	-13.4	n	-13.4	5.8	2	45	7.2
1000 lb	A-3	-28.5	y	7.0	5.8	2	45	7.2
1000 lb	A-4	-1.7	y	7.4	5.8	2	45	7.2
1000 lb	A-7	-41.6	y	6.5	5.8	2	45	7.2
1000 lb	A-8	-50.1	y	6.7	5.8	2	45	7.2
1000 lb	A-9	-60.1	y	12.2	5.8	2	45	7.2
1000 lb	A-10	-11.3	y	7.8	5.8	2	45	7.2
1000 lb	A-14	-23.7	y	8.4	5.8	2	45	7.2
1000 lb	A-17	-46.2	y	9.9	5.8	2	45	7.2
1000 lb	A-18	-6.5	y	7.2	5.8	2	45	7.2
1000 lb	A-20	-0.3	y	7.8	5.8	2	45	7.2

#### Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

7/19/2018

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

## Calibration Certificate of Mass

**Calibration Date:** July 17, 2018

**Certificate Number:** 201-063-2

**Submitted By:** FSCP Area 20  
944 N 20th Rd  
Unadilla, NE 68454

**Point of Contact:** Kurt Wenninghoff  
Ph. 402-471-3422  
**email:** [kurt.wenninghoff@nebraska.gov](mailto:kurt.wenninghoff@nebraska.gov)  
**PO Number:** N/A

**Test Item(s):** (1)-31 lb weight kit  
**Serial Number(s):** WM-289-4  
**Manufacture:** Rice Lake  
**Condition:** Fair (significant wear)

**Artifact(s) Description:**

**Date Received:** July 13, 2018

**ID / Asset Number:** N/A  
**Class Specification:** NIST Class F  
**Material:** SS & AL

**Reference Standards Used:**

NSL lb standards

**Procedure Used:**

NIST HB 6969, SOP 8

**Metrologist:**

JPL

**Equipment Used:**

Sartorius CC10000S    Mettler AT 106  
Sartorius CC 1201    Sartorius CCE6

**Environmental Cond.**    **Temp:** 22.8 °C    **Pressure:** 766.572 mmHg    **Relative Humidity:** 49 %

**Pertinent Information**

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

**Traceability Statement**

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

**Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: July 17, 2018

Certificate Number: 201-063-2

**Calibration Results**

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
5 lb	1	-0.082	n	-0.082	0.028	2	0.23	7.84
5 lb	2	-0.108	n	-0.108	0.028	2	0.23	7.84
5 lb	3	-0.102	n	-0.102	0.028	2	0.23	7.84
5 lb	4	-0.044	n	-0.044	0.028	2	0.23	7.84
5 lb	5	-0.060	n	-0.060	0.028	2	0.23	7.84
1 lb	1	0.0092	n	0.0092	0.0083	2	0.07	7.84
1 lb	2	0.0324	n	0.0324	0.0083	2	0.07	7.84
1 lb	3	-0.0244	n	-0.0244	0.0083	2	0.07	7.84
1 lb	4	0.0015	n	0.0015	0.0083	2	0.07	7.84
1 lb	5	-0.0279	n	-0.0279	0.0083	2	0.07	7.84
0.2 lb		0.0083	n	0.0083	0.0022	2	0.018	7.84
0.2 lb	*	0.0081	n	0.0081	0.0022	2	0.018	7.84
0.1 lb		0.0041	n	0.0041	0.0011	2	0.0091	7.84
0.05 lb		0.00168	n	0.00168	0.00054	2	0.0045	7.84
0.02 lb		0.00034	n	0.00034	0.00022	2	0.0018	7.84
0.02 lb	*	0.00028	n	0.00028	0.00022	2	0.0018	7.84
0.01 lb		0.00037	n	0.00037	0.00018	2	0.0015	7.84
0.005 lb		0.00053	n	0.00053	0.00015	2	0.0012	2.7
0.002 lb		0.00002	n	0.00002	0.00011	2	0.00087	2.7
0.002 lb	*	0.00005	n	0.00005	0.00011	2	0.00087	2.7
0.001 lb		0.000204	n	0.000204	0.000083	2	0.0007	2.7
8 oz		0.0010	n	0.0010	0.0054	2	0.045	7.84
4 oz		-0.0001	n	-0.0001	0.0028	2	0.023	7.84
2 oz		-0.0027	n	-0.0027	0.0013	2	0.011	7.84
1 oz		0.00168	n	0.00168	0.00064	2	0.0054	7.84
1/2 oz		0.00042	n	0.00042	0.00034	2	0.0028	7.84
1/4 oz		-0.00116	n	-0.00116	0.00021	2	0.0017	7.84
1/8 oz		0.00008	n	0.00008	0.00016	2	0.0013	7.84

**Conversion Factors**

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly

**Joel P. Lavicky Metrologist**

**7/18/2018**

**Date of Issue**

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

## Calibration Certificate of Mass

**Calibration Date:** July 9, 2018

**Certificate Number:** 2018-063-3

**Submitted By:** FSCP Area 20  
944 N 20th Rd  
Unadilla, NE 68454

**Point of Contact:** Kurt Wenninghoff  
Ph. 402-471-3422  
**email:** [kurt.wenninghoff@nebraska.gov](mailto:kurt.wenninghoff@nebraska.gov)  
**PO Number:** N/A

**Test Item(s):** 1-Metric weight kit  
**Serial Number(s):** WM-2D86  
**Manufacture:** Tromner  
**Condition:** Good (some wear)

**Artifact(s) Description:**

**Date Received:** July 13, 2018

**ID / Asset Number:** N/A  
**Class Specification:** NIST Class F  
**Material:** SS

**Reference Standards Used:**

**Procedure Used:**

**Equipment Used:**

OPI & /Den Metric  
Volland-1707

NIST HB 6969, SOP 8  
**Metrologist:**  
JPL

Sartorius CC 1201 Sartorius CCE6  
Mettler AT 106

**Environmental Cond.** Temp: 22.8 °C Pressure: 766.572 mmHg Relative Humidity: 49 %

**Pertinent Information**

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

**Traceability Statement**

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

**Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: July 9, 2018

Certificate Number: 2018-063-3

**Calibration Results**

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
1 kg	1	-0.003	n	-0.003	0.012	2	0.1	7.84
500 g	2	-0.0100	n	-0.0100	0.0083	2	0.07	7.84
200 g	3	0.0017	n	0.0017	0.0048	2	0.04	7.84
200 g	4	-0.0059	n	-0.0059	0.0048	2	0.04	7.84
100 g		0.0116	n	0.0116	0.0024	2	0.02	7.84
50 g		-0.0064	n	-0.0064	0.0012	2	0.01	7.84
20 g		0.00100	n	0.00100	0.00048	2	0.004	7.84
20 g	*	0.00123	n	0.00123	0.00048	2	0.004	7.84
10 g		-0.00107	n	-0.00107	0.00024	2	0.002	7.84
5 g		0.00002	n	0.00002	0.00018	2	0.0015	7.84
2 g		0.00050	n	0.00050	0.00013	2	0.0011	7.84
2 g	*	0.00052	n	0.00052	0.00013	2	0.0011	7.84
1 g		-0.00028	n	-0.00028	0.00011	2	0.0009	7.84
500 mg		0.000292	n	0.000292	0.000086	2	0.00072	7.95
200 mg		0.000325	n	0.000325	0.000064	2	0.00054	7.95
200 mg	*	0.000281	n	0.000281	0.000064	2	0.00054	7.95
100 mg		-0.000068	n	-0.000068	0.000051	2	0.00043	7.95

**Conversion Factors**

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly

Joel P. Lavicky Metrologist

7/18/2018

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.