

Agricultural Laboratory  
6531 SE Forbes Ave, Suite B  
Topeka, Kansas 66619  
(785) 296-7020



Office of the Secretary  
900 SW Jackson, Room 456  
Topeka, Kansas 66612  
(785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Test Date: 12/16/2015

Test No.: K15194-1.1

## Kansas Metrology Laboratory Certificate of Calibration



Nebraska Department Of Agriculture  
Food Safety & Consumer Protection  
301 Centennial Mall South  
Lincoln NE 68509

Manufacturer: Troemner

S/N: WM-G89-5

Number of Pieces: 11 of 23 total

Nominal Mass	Weight's Markings	Assumed Density (g/cm <sup>3</sup> )	Conventional Mass As Found (g)	Conventional Mass As Left (g)	Expanded Uncertainty ± (mg)	In Tolerance Adjusted Rejected
300 g	300g	7.84	300.001857	300.001857	0.098	In Tolerance
200 g	200g	7.84	200.001968	200.001968	0.093	In Tolerance
100 g	100g	7.84	100.001147	100.001147	0.018	In Tolerance
50 g	50g	7.84	50.000801	50.000801	0.011	In Tolerance
30 g	30g	7.84	30.0002951	30.0002951	0.0065	In Tolerance
20 g	20	7.84	20.0001523	20.0001523	0.0055	In Tolerance
10 g	10	7.84	10.0002913	10.0002913	0.0050	In Tolerance
5 g	5	7.84	5.0002288	5.0002288	0.0024	In Tolerance
3 g	3	7.84	3.0002805	3.0002805	0.0031	In Tolerance
2 g	2	7.84	2.0001037	2.0001037	0.0024	In Tolerance
1 g	1	7.84	1.0000237	1.0000237	0.0013	In Tolerance

This document certifies the above mentioned artifacts were compared to the Standards of the State of Kansas which are traceable to the National Institute of Standards and Technology. The conventional mass is the weight in normal air (1.2 mg/cm<sup>3</sup>) at 20 °C versus the reference density of 8.0 g/cm<sup>3</sup>. Calibration of listed items was performed according to NISTIR 6969, SOP 4 (Double Substitution) and/or NISTIR 5672, SOP 5 (3-1).

Tolerances were evaluated to ASTM Class 4. Surface finish and magnetism were not evaluated as it is assumed to be done by the manufacturer.

**Uncertainty Statement:**

The combined standard uncertainty includes the standard uncertainty reported for the standards, tare weights, the standard uncertainty for the measurement process, the standard uncertainty for air buoyancy corrections as stated in OIML R111-1 [2004E] eq. C.6.3-1, and a component of uncertainty to account for any observed deviations (Bias) from NIST (National Institute of Standards and Technology) values that are less than surveillance limits. Factors not considered in the evaluation: magnetism, weights are considered to meet magnetism specifications unless measurement aberrations are noted, balance eccentricity and linearity, these factors are considered as a part of the measurement assurance process when using a check standard with adequate degrees of freedom. The combined uncertainty is multiplied by the coverage factor ( $k$ -value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The coverage factor reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is also consistent with and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

**Uncertainty Analysis:**

Nominal	$S_p$	$U_S (k=1)$	$U_{tare} (k=1)$	$U_{Air\ Buoyancy\ Eq.}$	$P_{air}$	Procedure
300 g	0.0408	0.0172	No Tare	0.00209	1.14937	SOP 5
200 g	0.0408	0.0121	No Tare	0.00111	1.15063	SOP 5
100 g	0.00151	0.00907	No Tare	0.000610	1.15017	SOP 5
50 g	0.00238	0.00470	No Tare	0.000439	1.14961	SOP 5
30 g	0.00109	0.00302	No Tare	0.000225	1.15105	SOP 5
20 g	0.00158	0.00222	No Tare	0.000132	1.15194	SOP 5
10 g	0.00160	0.00183	No Tare	0.0000629	1.15225	SOP 5
5 g	0.000695	0.000970	No Tare	0.0000352	1.15259	SOP 5
3 g	0.00125	0.000650	No Tare	0.0000223	1.15217	SOP 5
2 g	0.00102	0.000495	No Tare	0.0000173	1.15077	SOP 5
1 g	0.000459	0.000455	No Tare	-0.00000774	1.16361	SOP 5

All values listed as a component of the overall uncertainty are in units of milligrams (mg) or (mg/cm<sup>3</sup>).

**Traceability Statement:**

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

**Condition of Item(s) Submitted for Testing:** Minor wear.  
**Treatment of Item(s) before Testing:** Item(s) were tested as found.  
**Documentary Standards:** NIST Handbook 105 Series, NISTIR 6969, SOP 4, NISTIR 5672, SOP 5, & ASTM E 617-13 or OIML R111-1  
**Item(s) Received on:** 12/14/2015  
**Item(s) Acclimated:** 12/14/2015 1:14:00 PM

**Environmental Conditions:**

Temperature	Barometric Pressure	Relative Humidity
20.2 °C	733.04 mmHg	43.0 %

Values are averages recorded over the duration of testing



**Keith Arkenberg , Metrologist**

12/23/2015

Date

KML Software Version: 8.3

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.

Agricultural Laboratory  
6531 SE Forbes Ave, Suite B  
Topeka, Kansas 66619  
(785) 296-7020



Office of the Secretary  
900 SW Jackson, Room 456  
Topeka, Kansas 66612  
(785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Test Date: 12/18/2015

Test No.: K15194-1.2

## Kansas Metrology Laboratory Certificate of Calibration



Nebraska Department Of Agriculture  
Food Safety & Consumer Protection  
301 Centennial Mall South  
Lincoln NE 68509

Manufacturer: Troemner  
S/N: WM-G89-10  
Number of Pieces: 12 of 23 total

Nominal Mass	Weight's Markings	Assumed Density (g/cm <sup>3</sup> )	Conventional Mass As Found (g)	Conventional Mass As Left (g)	Expanded Uncertainty ± (mg)	In Tolerance Adjusted Rejected
500 mg	500 mg	7.84	0.5000309	0.5000309	0.0012	In Tolerance
200 mg	200 mg	7.84	0.19996809	0.19996809	0.00066	In Tolerance
200 mg	200 mg ●	7.84	0.19997355	0.19997355	0.00066	In Tolerance
100 mg	100 mg	7.84	0.09993273	0.09993273	0.00073	In Tolerance
50 mg	50	7.84	0.04999602	0.04999602	0.00042	In Tolerance
20 mg	20	2.7	0.02001466	0.02001466	0.00028	In Tolerance
20 mg	20 ●	2.7	0.02001195	0.02001195	0.00028	In Tolerance
10 mg	10	2.7	0.00999644	0.00999644	0.00039	In Tolerance
5 mg	5	2.7	0.00500755	0.00500755	0.00034	In Tolerance
2 mg	2	2.7	0.00200626	0.00200626	0.00031	In Tolerance
2 mg	2 ●	2.7	0.00199883	0.00199883	0.00031	In Tolerance
1 mg	1	2.7	0.00100439	0.00100439	0.00039	In Tolerance

This document certifies the above mentioned artifacts were compared to the Standards of the State of Kansas which are traceable to the National Institute of Standards and Technology. The conventional mass is the weight in normal air (1.2 mg/cm<sup>3</sup>) at 20 °C versus the reference density of 8.0 g/cm<sup>3</sup>. Calibration of listed items was performed according to NISTIR 6969, SOP 4 (Double Substitution) and/or NISTIR 5672, SOP 5 (3-1).

Tolerances were evaluated to ASTM Class 4. Surface finish and magnetism were not evaluated as it is assumed to be done by the manufacturer.

**Uncertainty Statement:**

The combined standard uncertainty includes the standard uncertainty reported for the standards, tare weights, the standard uncertainty for the measurement process, the standard uncertainty for air buoyancy corrections as stated in OIML R111-1 [2004E] eq. C.6.3-1, and a component of uncertainty to account for any observed deviations (Bias) from NIST (National Institute of Standards and Technology) values that are less than surveillance limits. Factors not considered in the evaluation: magnetism, weights are considered to meet magnetism specifications unless measurement aberrations are noted, balance eccentricity and linearity, these factors are considered as a part of the measurement assurance process when using a check standard with adequate degrees of freedom. The combined uncertainty is multiplied by the coverage factor ( $k$ -value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The coverage factor reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is also consistent with and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

**Uncertainty Analysis:**

Nominal	$S_p$	$U_S (k=1)$	$U_{tare} (k=1)$	$U_{Air Buoyancy Eq.}$	$P_{air}$	Procedure
500 mg	0.000503	0.000255	No Tare	0.0000206	1.16252	SOP 5
200 mg	0.000284	0.000150	No Tare	0.00000821	1.16301	SOP 5
200 mg	0.000284	0.000150	No Tare	0.00000820	1.16395	SOP 5
100 mg	0.000324	0.000150	No Tare	0.00000411	1.16324	SOP 5
50 mg	0.000180	0.0000950	No Tare	0.00000206	1.16283	SOP 5
20 mg	0.000117	0.0000650	No Tare	0.0000218	1.16212	SOP 5
20 mg	0.000117	0.0000650	No Tare	0.0000218	1.16284	SOP 5
10 mg	0.000179	0.0000700	No Tare	0.0000110	1.16307	SOP 5
5 mg	0.000149	0.0000550	No Tare	0.00000549	1.16241	SOP 5
2 mg	0.000140	0.0000550	No Tare	0.00000220	1.16399	SOP 5
2 mg	0.000140	0.0000550	No Tare	0.00000220	1.16367	SOP 5
1 mg	0.000177	0.0000650	No Tare	0.00000110	1.16274	SOP 5

All values listed as a component of the overall uncertainty are in units of milligrams (mg) or (mg/cm<sup>3</sup>).

**Traceability Statement:**

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

**Condition of Item(s) Submitted for Testing:** Minor wear.

**Treatment of Item(s) before Testing:** Item(s) were tested as found.

**Documentary Standards:** NIST Handbook 105 Series, NISTIR 6969, SOP 4, NISTIR 5672, SOP 5, & ASTM E 617-13 or OIML R111-1

**Item(s) Received on:** 12/14/2015

**Item(s) Acclimated:** 12/14/2015 1:14:00 PM

**Environmental Conditions:**

Temperature	Barometric Pressure	Relative Humidity
20.3 °C	737.33 mmHg	43.2 %

Values are averages recorded over the duration of testing



**Keith Arkenberg, Metrologist**

12/23/2015

Date

KML Software Version: 8.3

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.



Agricultural Laboratory  
6531 SE Forbes Ave, Suite B  
Topeka, Kansas 66619  
(785) 296-7020



Office of the Secretary  
900 SW Jackson, Room 456  
Topeka, Kansas 66612  
(785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Test Date: 12/21/2015

Test No.: K15194-1.0

## Kansas Metrology Laboratory Certificate of Calibration



Nebraska Department Of Agriculture  
Food Safety & Consumer Protection  
301 Centennial Mall South  
Lincoln NE 68509

Manufacturer: Troemner

S/N: 5

Number of Pieces: 1

Nominal Mass	Weight's Markings	Assumed Density (g/cm <sup>3</sup> )	Conventional Mass As Found (g)	Conventional Mass As Left (g)	Expanded Uncertainty ± (mg)	In Tolerance Adjusted Rejected
4 kg	4 kg 5	7.84	4000.0611	4000.0611	3.7	In Tolerance

This document certifies the above mentioned artifacts were compared to the Standards of the State of Kansas which are traceable to the National Institute of Standards and Technology. The conventional mass is the weight in normal air (1.2 mg/cm<sup>3</sup>) at 20 °C versus the reference density of 8.0 g/cm<sup>3</sup>. Calibration of listed items was performed according to NISTIR 6969, SOP 4 (Double Substitution) and/or NISTIR 5672, SOP 5 (3-1).

Tolerances were evaluated to ASTM Class 4. Surface finish and magnetism were not evaluated as it is assumed to be done by the manufacturer.

**Uncertainty Statement:**

The combined standard uncertainty includes the standard uncertainty reported for the standards, tare weights, the standard uncertainty for the measurement process, the standard uncertainty for air buoyancy corrections as stated in OIML R111-1 [2004E] eq. C.6.3-1, and a component of uncertainty to account for any observed deviations (Bias) from NIST (National Institute of Standards and Technology) values that are less than surveillance limits. Factors not considered in the evaluation: magnetism, weights are considered to meet magnetism specifications unless measurement aberrations are noted, balance eccentricity and linearity, these factors are considered as a part of the measurement assurance process when using a check standard with adequate degrees of freedom. The combined uncertainty is multiplied by the coverage factor (*k*-value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The coverage factor reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is also consistent with and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

### Uncertainty Analysis:

Nominal	$S_p$	$u_S (k=1)$	$u_{tare} (k=1)$	$u_{Air} \text{ Buoyancy Eq.}$	$\rho_{air}$	Procedure
4 kg	1.68	0.310	0.0535	0.0132	1.15353	SOP 4

All values listed as a component of the overall uncertainty are in units of milligrams (mg) or (mg/cm<sup>3</sup>).

**Traceability Statement:**

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

**Condition of Item(s) Submitted for Testing:** Minor wear.

**Treatment of Item(s) before Testing:** Item(s) were tested as found.

**Documentary Standards:** NIST Handbook 105 Series, NISTIR 6969, SOP 4, NISTIR 5672, SOP 5, & ASTM E 617-13 or OIML R111-1

Item(s) Received on: 12/14/2015

Item(s) Acclimated: 12/14/2015 1:14:00 PM

**Environmental Conditions:**

Temperature	Barometric Pressure	Relative Humidity
20.5 °C	731.72 mmHg	44.9 %

Values are averages recorded over the duration of testing



12/23/2015

Date \_\_\_\_\_

KML Software Version: 8.3

This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.

Agricultural Laboratory  
6531 SE Forbes Ave, Suite B  
Topeka, Kansas 66619  
(785) 296-7020



Office of the Secretary  
900 SW Jackson, Room 456  
Topeka, Kansas 66612  
(785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Expires on: 12/16/2016

## Kansas Metrology Laboratory Calibration Report

Report Number: K15194

Submitted by:

Nebraska Department Of Agriculture  
Food Safety & Consumer Protection  
301 Centennial Mall South  
Lincoln NE 68509

Submitted on: 12/14/2015

Item(s)		
Tested	Adjusted	Rejected
70	6	2

Quantity	Nominal Mass	Type
20	25 lb	Weight(s)
2	15 lb	Weight(s)
37	2 lb to 0.001 lb 8 oz to 1/16 oz	Weight Kit
11	0.2 lb to 0.001 lb	Weight Kit

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 10S-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance $\pm$	Expanded Uncertainty (U), (k=2), $\pm$	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
25 lb	NE-100	<b>11338.88 g</b>	1.1 g	0.17 g	<b>11339.89 g</b>	<b>Adjusted</b>
25 lb	NE-81	<b>11338.90 g</b>	1.1 g	0.17 g	<b>11338.90 g</b>	In Tolerance
25 lb	NE-82	<b>11340.32 g</b>	1.1 g	0.17 g	<b>11340.32 g</b>	In Tolerance
25 lb	NE-83	<b>11339.23 g</b>	1.1 g	0.17 g	<b>11339.23 g</b>	In Tolerance
25 lb	NE-94	<b>11339.30 g</b>	1.1 g	0.17 g	<b>11339.30 g</b>	In Tolerance
25 lb	NE-99	<b>11339.73 g</b>	1.1 g	0.17 g	<b>11339.73 g</b>	In Tolerance
25 lb	WM25-106	<b>11338.54 g</b>	1.1 g	0.17 g	<b>11339.83 g</b>	<b>Adjusted</b>
25 lb	WM25-113	<b>11339.17 g</b>	1.1 g	0.17 g	<b>11339.17 g</b>	In Tolerance
25 lb	WM25-27	<b>11339.47 g</b>	1.1 g	0.17 g	<b>11339.47 g</b>	In Tolerance
25 lb	WM25-29	<b>11339.26 g</b>	1.1 g	0.17 g	<b>11339.26 g</b>	In Tolerance
25 lb	WM25-65	<b>11339.06 g</b>	1.1 g	0.17 g	<b>11339.06 g</b>	In Tolerance
25 lb	WM25-66	<b>11338.90 g</b>	1.1 g	0.17 g	<b>11338.90 g</b>	In Tolerance
25 lb	WM25-67	<b>11339.32 g</b>	1.1 g	0.17 g	<b>11339.32 g</b>	In Tolerance
25 lb	WM25-68	<b>11338.99 g</b>	1.1 g	0.17 g	<b>11338.99 g</b>	In Tolerance
25 lb	WM25-69	<b>11338.99 g</b>	1.1 g	0.17 g	<b>11338.99 g</b>	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb  
28.349523125 g = 1 oz

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance $\pm$	Expanded Uncertainty (U), (k=2), $\pm$	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
25 lb	WM25-70	<b>11338.27 g</b>	1.1 g	0.17 g	<b>11339.89 g</b>	<b>Adjusted</b>
25 lb	WM25-71	<b>11339.18 g</b>	1.1 g	0.17 g	<b>11339.18 g</b>	In Tolerance
25 lb	WM25-72	<b>11339.50 g</b>	1.1 g	0.17 g	<b>11339.50 g</b>	In Tolerance
25 lb	WM25-73	<b>11340.05 g</b>	1.1 g	0.17 g	<b>11340.05 g</b>	In Tolerance
25 lb	WM25-74	<b>11340.24 g</b>	1.1 g	0.17 g	<b>11340.24 g</b>	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb

28.349523125 g = 1 oz

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance $\pm$	Expanded Uncertainty (U), (k=2), $\pm$	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
15 lb	WM15-13	<b>6805.732 g</b>	0.68 g	0.090 g	<b>6803.852 g</b>	<b>Adjusted</b>
15 lb	WM15-14	<b>6804.712 g</b>	0.68 g	0.090 g	<b>6803.932 g</b>	<b>Adjusted</b>

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb

28.349523125 g = 1 oz



The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance $\pm$	Expanded Uncertainty (U), (k=2), $\pm$	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
2 lb	12A9 1	907.105 g	0.091 g	0.011 g	907.105 g	In Tolerance
2 lb	12A9 2	907.145 g	0.091 g	0.011 g	907.145 g	In Tolerance
2 lb	12A9 3	907.151 g	0.091 g	0.011 g	907.151 g	In Tolerance
2 lb	12A9 4	907.140 g	0.091 g	0.011 g	907.140 g	In Tolerance
2 lb	12A9 5	907.168 g	0.091 g	0.011 g	907.168 g	In Tolerance
2 lb	12A9 6	907.178 g	0.091 g	0.011 g	907.178 g	In Tolerance
2 lb	12A9 7	907.193 g	0.091 g	0.011 g	907.193 g	In Tolerance
2 lb	12A9 8	907.145 g	0.091 g	0.011 g	907.145 g	In Tolerance
2 lb	12A9 9	907.172 g	0.091 g	0.011 g	907.172 g	In Tolerance
2 lb	12A9 10	907.193 g	0.091 g	0.011 g	907.193 g	In Tolerance
2 lb	12A9 11	907.159 g	0.091 g	0.011 g	907.159 g	In Tolerance
2 lb	12A9 12	907.161 g	0.091 g	0.011 g	907.161 g	In Tolerance
2 lb	12A9 13	907.171 g	0.091 g	0.011 g	907.171 g	In Tolerance
2 lb	12A9 14	907.204 g	0.091 g	0.011 g	907.204 g	In Tolerance
1 lb	12A9 1	453.5669 g	0.070 g	0.0084 g	453.5669 g	In Tolerance
1 lb	12A9 2	453.5629 g	0.070 g	0.0084 g	453.5629 g	In Tolerance
0.3 lb	12A9	136.0581 g	0.027 g	0.0032 g	136.0581 g	In Tolerance
0.2 lb	12A9	90.7044 g	0.018 g	0.0021 g	90.7044 g	In Tolerance
0.1 lb	12A9	45.3572 g	0.0091 g	0.0011 g	45.3572 g	In Tolerance
0.05 lb	12A9	22.67557 g	0.0045 g	0.00055 g	22.67557 g	<b>Rejected</b>
0.03 lb	12A9	13.60567 g	0.0027 g	0.00032 g	13.60567 g	In Tolerance
0.02 lb	12A9	9.07196 g	0.0018 g	0.00022 g	9.07196 g	In Tolerance
0.01 lb	12A9	4.53523 g	0.0015 g	0.00018 g	4.53523 g	In Tolerance
0.005 lb	12A9	2.26790 g	0.0012 g	0.00015 g	2.26790 g	In Tolerance
0.003 lb	12A9	1.36130 g	0.00099 g	0.00012 g	1.36130 g	In Tolerance
0.002 lb	12A9	0.90739 g	0.00087 g	0.00011 g	0.90739 g	In Tolerance
0.001 lb	12A9	0.453542 g	0.00070 g	0.000094 g	0.453542 g	In Tolerance
0.001 lb	12A9 •	0.453372 g	0.00070 g	0.000094 g	0.453372 g	In Tolerance
8 oz	12A9	226.7505 g	0.045 g	0.0053 g	226.7915 g	<b>Adjusted</b>
4 oz	12A9	113.3885 g	0.023 g	0.0028 g	113.3885 g	In Tolerance
2 oz	12A9	56.6967 g	0.011 g	0.0013 g	56.6967 g	In Tolerance
1 oz	12A9	28.34921 g	0.0054 g	0.00065 g	28.34921 g	In Tolerance
1/2 oz	12A9	14.17273 g	0.0028 g	0.00033 g	14.17273 g	In Tolerance
1/4 oz	12A9	7.08714 g	0.0017 g	0.00020 g	7.08714 g	In Tolerance
1/8 oz	12A9	3.54262 g	0.0013 g	0.00016 g	3.54262 g	In Tolerance
1/16 oz	12A9	1.77095 g	0.0011 g	0.00014 g	1.77095 g	In Tolerance
1/16 oz	12A9 •	1.77101 g	0.0011 g	0.00014 g	1.77101 g	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb  
28.349523125 g = 1 oz

The calibration of items is performed according to NISTIR 6969, SOP 8. Tolerances are applied from NISTHB 105-1.

Nominal Mass	Serial Number	Conventional Mass as Found	Tolerance $\pm$	Expanded Uncertainty (U), (k=2), $\pm$	Conventional Mass as Left	Adjusted/ In Tolerance/ Rejected
0.2 lb	N-99-A	<b>90.7245 g</b>	0.018 g	0.0021 g	<b>90.7245 g</b>	In Tolerance
0.2 lb	N-99-A •	<b>90.7248 g</b>	0.018 g	0.0021 g	<b>90.7248 g</b>	In Tolerance
0.1 lb	N-99-A	<b>45.3615 g</b>	0.0091 g	0.0011 g	<b>45.3615 g</b>	In Tolerance
0.05 lb	N-99-A	<b>22.67855 g</b>	0.0045 g	0.00055 g	<b>22.67855 g</b>	In Tolerance
0.02 lb	N-99-A	<b>9.07150 g</b>	0.0018 g	0.00022 g	<b>9.07150 g</b>	In Tolerance
0.02 lb	N-99-A •	<b>9.07169 g</b>	0.0018 g	0.00022 g	<b>9.07169 g</b>	In Tolerance
0.01 lb	N-99-A	<b>4.53654 g</b>	0.0015 g	0.00018 g	<b>4.53654 g</b>	In Tolerance
0.005 lb	N-99-A	<b>2.26725 g</b>	0.0012 g	0.00015 g	<b>2.26725 g</b>	In Tolerance
0.002 lb	N-99-A	<b>0.90722 g</b>	0.00087 g	0.00011 g	<b>0.90722 g</b>	In Tolerance
0.002 lb	N-99-A •	<b>0.90637 g</b>	0.00087 g	0.00011 g	<b>0.90637 g</b>	<b>Rejected</b>
0.001 lb	N-99-A	<b>0.454082 g</b>	0.00070 g	0.000094 g	<b>0.454082 g</b>	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

453.59237 g = 1 lb

28.349523125 g = 1 oz

**Uncertainty Statement:**

The combined standard uncertainty includes the standard uncertainty reported for the standard, the standard uncertainty for the measurement process, the standard uncertainty for any uncorrected errors associated with buoyancy corrections (applies to mass values only), the standard uncertainty for any uncorrected errors associated with temperature correction (applies to length and volume values only), and a component of uncertainty to account for any observed deviations from NIST (The National Institute of Standards and Technology) values that are less than surveillance limits. The combined standard uncertainty is multiplied by a coverage factor of 2 to give an expanded uncertainty, which defines an interval having a level of confidence of approximately 95 percent. The expanded uncertainty presented in this report is consistent with the 1993 ISO Guide to the Expression of Uncertainty in Measurement and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

**Traceability Statement:**

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

**Condition of Item(s) Submitted for Testing:**

Minor wear.

**Treatment of Item(s) before Testing:**

Item(s) were tested as found.

**Documentary Standards:**


NIST Handbook 10S Series  
NISTIR 6969: SOP 8, SOP 4, and/or SOP 7  
OR  
ASTM E 617-13 or OIML R 111-1 2004(E)

**Environmental Conditions:**

Temperature:	19.9 °C
Barometric Pressure:	735.43 mmHg
Relative Humidity:	44.9 %

Test Date: 12/17/2015

Due Date: 12/16/2016 -Per state statute K.S.A. 83-304(a).

  
**Keith Arkenberg , Metrologist**

12/23/2015



This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.



Agricultural Laboratory  
6531 SE Forbes Ave, Suite B  
Topeka, Kansas 66619  
(785) 296-7020



Office of the Secretary  
900 SW Jackson, Room 456  
Topeka, Kansas 66612  
(785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

Expires on: 12/16/2016

## Kansas Metrology Laboratory Calibration Report

Report Number: K15194-TM

Submitted by:

Nebraska Department Of Agriculture  
Food Safety & Consumer Protection  
301 Centennial Mall South  
Lincoln NE 68509

Submitted on: 12/14/2015

Item(s)		
Tested	Adjusted	Rejected
5	3	0
Quantity	Nominal Volume	Type
3	5 gal	Bottom Drop Test Measure "To Deliver"
2	5 gal	Handheld Test Measure "To Deliver"

The calibration of items is performed according to NISTIR 7383, SOP 19 Volume Transfer. Tolerances are applied from NISTHB 105-3. The volume applies when a 10 second drain is observed for 5 gallon hand held test measures. For 5 gallon bottom drop test measures and provers a 30 second drain applies. The drain time starts when the cessation of the main flow is observed.

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (/°F)	Volume as Found @ 60 °F	Tolerance ±	Expanded Uncertainty (U), (k=2.02), ±	Volume as Left @ 60 °F	Adjusted/ In Tolerance/ Rejected
5 gal	0236	Stainless Steel	0.0000265	5.00317 gal	0.0025 gal	0.00084 gal	4.99971 gal	Adjusted
5 gal	0237	Stainless Steel	0.0000265	5.00404 gal	0.0025 gal	0.00084 gal	4.99928 gal	Adjusted
5 gal	0238	Stainless Steel	0.0000265	5.00395 gal	0.0025 gal	0.00085 gal	5.00092 gal	Adjusted
5 gal	87276	Stainless Steel	0.0000265	4.99971 gal	0.0025 gal	0.00082 gal	4.99971 gal	In Tolerance
5 gal	87280	Stainless Steel	0.0000265	4.99941 gal	0.0025 gal	0.00082 gal	4.99941 gal	In Tolerance

The data in the above table of this report only applies to those items specifically listed on this report.

1 m³=1 000 L=264.1720 gal



## Temperature Corrections

Item	Temperature °F	in <sup>3</sup>
Temperature Correction for 5 gal Stainless Steel Test Measure (CCE= 0.0000265/°F)	-20	-1.72
	-15	-1.61
	-10	-1.50
	-5	-1.40
	0	-1.29
	5	-1.18
	10	-1.07
	15	-0.97
	20	-0.86
	25	-0.75
	30	-0.64
	35	-0.54
	40	-0.43
	45	-0.32
	50	-0.21
	55	-0.11
	60	0.00
	65	0.11
	70	0.21
	75	0.32
	80	0.43
	85	0.54
	90	0.64
	95	0.75
	100	0.86
	105	0.97
	110	1.07
	115	1.18
	120	1.29

Item	Temperature °F	in <sup>3</sup>
Temperature Correction for 5 gal Low Carbon Steel Test Measure (CCE=0.0000186/°F)	-20	-1.72
	-15	-1.61
	-10	-1.50
	-5	-1.40
	0	-1.29
	5	-1.18
	10	-1.07
	15	-0.97
	20	-0.86
	25	-0.75
	30	-0.64
	35	-0.54
	40	-0.43
	45	-0.32
	50	-0.21
	55	-0.11
	60	0.00
	65	0.11
	70	0.21
	75	0.32
	80	0.43
	85	0.54
	90	0.64
	95	0.75
	100	0.86
	105	0.97
	110	1.07
	115	1.18
	120	1.29

CCE = Coefficient of Cubical Expansion

Expires on: 12/16/2016

# Kansas Metrology Laboratory

Report Number: K15194-TM

## Uncertainty Statement:

The combined standard uncertainty includes the standard uncertainty reported for the standards, the standard uncertainty for the measurement process, the standard uncertainty for the water density equation (Metrologia Tanaka, et al), the standard uncertainty for any uncorrected errors associated with temperature correction (applies to length and volume values only), the standard uncertainty for reading the meniscus (when applicable), the standard uncertainty for viscosity, and a component of uncertainty to account for any observed deviations from NIST (The National Institute of Standards and Technology) values that are less than surveillance limits. The combined standard uncertainty is multiplied by the coverage factor (k-value) reported to give an expanded uncertainty, which defines an interval having a level of confidence of 95.45 percent. The k-value reported is based on the effective degrees of freedom as outlined in JCGM 100:2008 section G.4. The expanded uncertainty presented in this report is consistent with the 1993 ISO Guide to the Expression of Uncertainty in Measurement and follows NISTIR 6969, SOP 29. The expanded uncertainty is not to be confused with a tolerance limit for the user during application.

## Traceability Statement:

The Kansas Metrology Laboratory Standards are traceable to the SI through NIST and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory test number identified above is the unique report number to be used in referencing measurement traceability for artifacts identified in this report only.

## Condition of Item(s) Submitted for Testing:

Minor wear.

## Treatment of Item(s) before Testing:

Item(s) were tested as found.

## Water Temperature at Time of Test:

70.14 °F

## Documentary Standards:

-NIST Handbook 105 Series

-NISTIR 7383, SOP 19

## Environmental Conditions:

Temperature:	20.96 °C
Barometric Pressure:	745.41 mmHg
Relative Humidity:	31.5 %

Test Date: 12/17/2015

Due Date: 12/16/2016 -Per state statute K.S.A. 83-304(a).



**Keith Arkenberg , Metrologist**

12/23/2015



This document does not represent or imply endorsement by the Kansas Metrology Laboratory, NIST, or any agency of the State and/or national governments. This document may not be reproduced, except in full, without the written permission of the Kansas Metrology Laboratory.



Traceable Report Number: 2402646  
 Contractor: STATE OF NEBRASKA  
 PO BOX 94757  
 LINCOLN, NE 68509-0757

ANSI/NCSL Z540-1-1994; Part 1 &amp; ISO/IEC 17025 Accredited

Purchase Order Number: 4024716842  
 Client: STATE OF NEBRASKA  
 PO BOX 94757  
 LINCOLN, NE 68509-0757

Date Received: 18 Feb 2016  
 Date Calibrated: 22 Feb 2016  
 Contractor Requested Recall Date: No Recall Requested  
 Temperature Range: 20.31 °C  
 Pressure Range: 735.44 mmHg  
 Relative Humidity Range: 52.73 %  
 Air Density Range: 1.1588 mg/cm<sup>3</sup>  
 NIST Certificate Number: 684/286541-15,684/284451-14  
 Although there are two NIST numbers, one or both may apply.  
 Tested By: 22  
 Procedure: Modified Substitution (WI05-0023)  
 Description of Weights: 0.05 lb Satin Finish Weight, NIST Class F, S/N 6GLA.



## Conventional Mass Corr.

Nominal Value	ID	As Found (mg)	As Found In Tol	As Left (mg)	As Left In Tol	Unc. (mg)	k	Tol.* (mg)	Balance Used	Standard Set Used	Assumed Density (g/cm <sup>3</sup> )
0.05 lb	6GLA.	1.43	Y	1.43	Y	0.57	2	4.5	1221Q	D564Q	7.84



This report contains data not covered by the NVLAP Accreditation if the box is checked.

Check with your local state agency for certification of compliance on Legal for Trade items.

The weight tolerance class is referenced in the Description of Weights. Unless otherwise noted, weights tested meet the requirements of the class.

\*The specifications for the weight classes can be found in NIST Handbook 105-1, ASTM E-617 or OIML R111.

Prepared By:

Rice Lake Weighing Systems

230 West Coleman Street, Rice Lake, WI 54868 • USA

TEL: 715-234-9171 • FAX: 715-234-6967 • [www.ricelake.com](http://www.ricelake.com)

An ISO 9001 registered company

Dated 23 Feb 2016

Dan Demers, Metrologist

Page 1 of 1



NVLAP Lab Code 105001-0

This report is not to be used to claim product endorsement by Rice Lake Weighing Systems, NVLAP or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory. Uncertainties are assigned using NIST SOP 29 in conformance with the GUM and include uncertainty components for reference standard, process, air buoyancy correction and bias/drift. Not included are magnetism, handling and use. The coverage factor *k* provides a level of confidence of approximately 95 percent.

PN 64784 11/16



Traceable Report Number: 2402646A  
 Contractor: STATE OF NEBRASKA  
 PO BOX 94757  
 LINCOLN, NE 68509-0757

ANSI/NCSL Z540-1-1994; Part 1 &amp; ISO/IEC 17025 Accredited

Purchase Order Number: 4024716842  
 Client: STATE OF NEBRASKA  
 PO BOX 94757  
 LINCOLN, NE 68509-0757

Date Received: 18 Feb 2016  
 Date Calibrated: 22 Feb 2016  
 Contractor Requested Recall Date: No Recall Requested

Temperature Range: 20.35 °C  
 Pressure Range: 735.44 mmHg

Relative Humidity Range: 54.36 %  
 Air Density Range: 1.1584 mg/cm<sup>3</sup>

NIST Certificate Number: 684/286541-15,684/284451-14  
 Although there are two NIST numbers, one or both may apply.

Tested By: 22  
 Procedure: Modified Substitution (WI05-0023)

Description of Weights: 0.002 lb Aluminum Dish Weight, NIST Class F, S/N 6GLC.



## Conventional Mass Corr.

Nominal Value	ID	As Found (mg)	As Found In Tol	As Left (mg)	As Left In Tol	Unc. (mg)	k	Tol.* (mg)	Balance Used	Standard Set Used	Assumed Density (g/cm <sup>3</sup> )
0.002 lb	6GLC.	0.30	Y	0.30	Y	0.11	2	0.87	638Q	D564Q	2.7

☐ This report contains data not covered by the NVLAP Accreditation if the box is checked.

Check with your local state agency for certification of compliance on Legal for Trade items.

The weight tolerance class is referenced in the Description of Weights. Unless otherwise noted, weights tested meet the requirements of the class.


\*The specifications for the weight classes can be found in NIST Handbook 105-1, ASTM E-617 or OIML R111.

Prepared By:

Rice Lake Weighing Systems

230 West Coleman Street, Rice Lake, WI 54868 • USA  
 TEL: 715-234-9171 • FAX: 715-234-6967 • [www.ricelake.com](http://www.ricelake.com)  
 An ISO 9001 registered company

Dated 23 Feb 2016

  
 Dan Demers, Metrologist

Page 1 of 1