

**Calibration Date:** 9/25/2017

**Certificate of Calibration  
of Volume Transfer**

**Certificate Number:** 2017-027-1

**Items Submitted:**

Quantity	Nominal Volume	Manufacturer	Type
2	5 gal	Seraphin	Test Measure

**Submitted By:** FSCP Area 90  
1118 Eldorado Dr.  
Omaha NE, 68154

**POC:** Brian Heskin  
402-496-0843

**Test Results**

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (1/°F)	As Found Volume Delivered @ 60 °F	As left Volume Delivered @ 60 °F	Uncertainty (U)	(k)
5 gal	40702A	SS	0.0000265	4.99774 gal	5.00011 gal	0.00061 gal	2.03
5 gal	40702B	SS	0.0000265	5.0012 gal	5.0012 gal	0.00061 gal	2.03

*The data in this report only applies to those items specifically listed on this report.*

Volume delivered at 60°F after a 30 second pour and 10 second drain for test measures. For provers and a 30 second drain time would apply.

**Conversion Factors:**

1 gal = 231 in<sup>3</sup>  
1 gal = 3.785 412 E-03 m<sup>3</sup>

**Traceability Statement:**

The artifact(s) described in this report have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this report is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this report.

**Uncertainty Statement:**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors. The combined standard uncertainty is multiplied by a coverage factor (k), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken.

**Pertinent Information:**

The artifact(s) listed above have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error.

**Condition of Item(s) Submitted for Calibration:**

Minor wear

**Laboratory Reference Standard Used:**

5 gal SP NE 1586

**Treatment of Item(s) before Calibration:**

Item(s) were tested as found

**Procedure Used:**

NISTIR 7383 (2017), SOP 19


**Environmental conditions at time of calibration:**

Temp °C	22.9	Humidity %	53.2
Pressure mmHg	763.27		

**Water temperature at time of calibration:**

67.97 °F

**Date Submitted:** 9/22/2017



Joel P. Lavicky, Metrologist

9/25/2017

Date:

*This document does not represent or imply endorsement by the State of Nebraska, The Nebraska Standards Laboratory or NIST. This document may not be reproduced, except in full, without the written permission of the Nebraska Standards Laboratory*

Calibration Date: 9/26/2017

**Certificate of Calibration  
of Volume Transfer**

Certificate Number: 2017-027-2

**Items Submitted:**

Quantity	Nominal Volume	Manufacturer	Type
3	5 gal	Seraphin	"Special" J Prover

Submitted By: FSCP Area 90  
1118 Eldorado Dr.  
Omaha NE, 68154

POC: Brian Heskin  
402-496-0843

**Test Results**

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (/°F)	As Found Volume Delivered @ 60 °F	As left Volume Delivered @ 60 °F	Uncertainty (U)	(k)
5 gal	05-40547-04	SS	0.0000265	4.99945 gal	4.99945 gal	0.00052 gal	2.03
5 gal	05-40547-05	SS	0.0000265	4.99902 gal	4.99902 gal	0.00052 gal	2.03
5 gal	05-40547-06	SS	0.0000265	4.99925 gal	4.99925 gal	0.00052 gal	2.03

The data in this report only applies to those items specifically listed on this report.

Volume delivered at 60°F after a 30 second pour and 10 second drain for test measures. For provers and a 30 second drain time would apply.

**Conversion Factors:**

1 gal = 231 in<sup>3</sup>  
1 gal = 3.785 412 E-03 m<sup>3</sup>

**Traceability Statement:**

The artifact(s) described in this report have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this report is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this report.

**Uncertainty Statement:**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken.

**Pertinent Information:**

The artifact(s) listed above have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error.

**Condition of Item(s) Submitted for Calibration:**

Minor wear

**Laboratory Reference Standard Used:**

5 gal SP NE 1586

**Treatment of Item(s) before Calibration:**

Item(s) were tested as found

**Procedure Used:**

NISTIR 7383 (2017), SOP 19


**Environmental conditions at time of calibration:**

Temp °C	21.1	Humidity %	55.4
Pressure mmHg	763.27		

**Water temperature at time of calibration:**

68.08 °F

Date Submitted: 9/22/2017

  
Joel P. Lavicky, Metrologist

9/26/2017

Date:

This document does not represent or imply endorsement by the State of Nebraska, The Nebraska Standards Laboratory or NIST. This document may not be reproduced, except in full, without the written permission of the Nebraska Standards Laboratory

## Calibration Certificate of Mass

**Calibration Date:** October 2, 2017

**Certificate Number:** 2017-027-3

**Submitted By:** FSCP Area 90  
1118 Eldorado Dr  
Omaha, NE

**Point of Contact:** Brian Heskin  
Ph. 402-672-2076  
**email:** [brian.heskin@nebraska.gov](mailto:brian.heskin@nebraska.gov)  
**PO Number:**

<b>Test Item(s):</b> 1-4 kg, 2-15 lb & 20-25 lb weights	<b>Artifact(s) Description:</b>	<b>Date Received:</b> September 22, 2017
<b>Serial Number(s):</b> See Next Page		<b>ID / Asset Number:</b> N/A
<b>Manufacture:</b> Tromner / Rice Lake		<b>Class Specification:</b> NIST Class F
<b>Condition:</b> Good (some wear)		<b>Material:</b> SS and CI

**Reference Standards Used:**

2kgd&2kg  
NEBR-STD-10&5  
NSL-25-1-25lb

**Procedure Used:**

NIST HB 6969, SOP 8

**Metrologist:**  
JPL

**Equipment Used:**

Sartorius CC10000S  
Mettler KA30-3

**Environmental Cond.**    **Temp:** 22.7 °C    **Pressure:** 767.08 mmHg    **Relative Humidity:** 52 %

**Pertinent Information**

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

**Traceability Statement**

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

**Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: October 2, 2017

Certificate Number: 2017-027-3


### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
4 kg	WM-7	0.000	n	0.000	0.048	2	0.4	7.94
15 lb	WM15-7	0.573	n	0.573	0.081	2	0.68	7.2
15 lb	WM15-8	0.254	n	0.254	0.081	2	0.68	7.2
25 lb	WM-D3	2.47	y	0.36	0.14	2	1.1	7.2
25 lb	WM-D4	1.42	y	0.75	0.14	2	1.1	7.2
25 lb	WM-D5	1.63	y	0.38	0.14	2	1.1	7.2
25 lb	WM-D6	1.66	y	0.69	0.14	2	1.1	7.2
25 lb	WM-D7	1.72	y	0.41	0.14	2	1.1	7.2
25 lb	WM-D8	2.04	y	0.69	0.14	2	1.1	7.2
25 lb	WM-D9	1.97	y	0.57	0.14	2	1.1	7.2
25 lb	WM-D10	3.36	y	0.52	0.14	2	1.1	7.2
25 lb	WM-D11	2.56	y	0.71	0.14	2	1.1	7.2
25 lb	WM-d12	2.31	y	0.38	0.14	2	1.1	7.2
25 lb	WM25-45	2.02	y	0.53	0.14	2	1.1	7.2
25 lb	WM25-46	1.58	y	0.46	0.14	2	1.1	7.2
25 lb	WM25-47	2.24	y	0.39	0.14	2	1.1	7.2
25 lb	WM25-88	1.78	y	-0.02	0.14	2	1.1	7.2
25 lb	WM25-89	2.00	y	0.22	0.14	2	1.1	7.2
25 lb	WM25-90	2.83	y	0.22	0.14	2	1.1	7.2
25 lb	WM25-91	1.85	y	0.33	0.14	2	1.1	7.2
25 lb	WM25-92	2.28	y	-0.18	0.14	2	1.1	7.2
25 lb	WM25-93	2.28	y	0.77	0.14	2	1.1	7.2
25 lb	WM25-94	2.18	y	0.41	0.14	2	1.1	7.2

#### Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

10/6/2017

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

## Calibration Certificate of Mass

**Calibration Date:** September 27, 2017

**Certificate Number:** 2017-027-4

**Submitted By:** FSCP Area 90  
1118 Eldorado Dr.  
Omaha, NE 68514

**Point of Contact:** Brian Heskin  
Ph. 402-496-0843  
**email:** [Brian.heskin@nebraska.gov](mailto:Brian.heskin@nebraska.gov)  
**PO Number:**

**Test Item(s):** 31 lb weight kit  
**Serial Number(s):** 10-OPI-9/WM-3G95  
**Manufacture:** Tromner/Rice Lake  
**Condition:** Good (some wear)

**Artifact(s) Description:**

**Date Received:** September 22, 2017

**ID / Asset Number:** N/A  
**Class Specification:** NIST Class F  
**Material:** Stainless Steel

**Reference Standards Used:**

NSL lb standards

**Procedure Used:**

NIST HB 6969, SOP 8

**Metrologist:**

JPL

**Equipment Used:**

Sartorius CCE6 Sartorius CC 1201  
Mettler AT 106

**Environmental Cond.** Temp: 22.25 °C Pressure: 771.65 mmHg Relative Humidity: 49.5 %

**Pertinent Information**

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

**Traceability Statement**

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

**Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.



Calibration Date: September 27, 2017

Certificate Number: 2017-027-4


### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
2 lb	1	-0.030	n	-0.030	0.011	2	0.091	7.84
2 lb	2	-0.020	n	-0.020	0.011	2	0.091	7.84
2 lb	3	-0.014	n	-0.014	0.011	2	0.091	7.84
2 lb	4	-0.016	n	-0.016	0.011	2	0.091	7.84
2 lb	5	0.003	n	0.003	0.011	2	0.091	7.84
2 lb	6	-0.004	n	-0.004	0.011	2	0.091	7.84
2 lb	7	0.016	n	0.016	0.011	2	0.091	7.84
2 lb	8	0.017	n	0.017	0.011	2	0.091	7.84
2 lb	9	-0.059	n	-0.059	0.011	2	0.091	7.84
2 lb	10	0.020	n	0.020	0.011	2	0.091	7.84
2 lb	11	0.030	n	0.030	0.011	2	0.091	7.84
2 lb	12	0.039	n	0.039	0.011	2	0.091	7.84
2 lb	13	0.027	n	0.027	0.011	2	0.091	7.84
2 lb	14	-0.002	n	-0.002	0.011	2	0.091	7.84
1 lb	15	-0.0114	n	-0.0114	0.0083	2	0.07	7.84
8 oz		-0.0205	n	-0.0205	0.0054	2	0.045	7.84
4 oz		0.0087	n	0.0087	0.0028	2	0.023	7.84
2 oz		0.0051	n	0.0051	0.0013	2	0.011	7.84
1 oz		-0.00413	n	-0.00413	0.00064	2	0.0054	7.84
1/2 oz		0.00123	n	0.00123	0.00034	2	0.0028	7.84
1/4 oz		0.00059	n	0.00059	0.00021	2	0.0017	7.84
1/8 oz		-0.00025	n	-0.00025	0.00016	2	0.0013	7.84
1/16 oz		0.00078	n	0.00078	0.00013	2	0.0011	7.84
1/16 oz	*	-0.00042	n	-0.00042	0.00013	2	0.0011	7.84
0.3 lb		0.0007	n	0.0007	0.0032	2	0.027	7.84
0.2 lb		0.0041	n	0.0041	0.0022	2	0.018	7.84
0.1 lb		0.0031	n	0.0031	0.0011	2	0.0091	7.84
0.05 lb		0.00130	n	0.00130	0.00054	2	0.0045	7.84
0.03 lb		0.00060	n	0.00060	0.00032	2	0.0027	7.84
0.02 lb		0.00053	n	0.00053	0.00022	2	0.0018	7.84
0.01 lb		0.00031	n	0.00031	0.00018	2	0.0015	7.84
0.005 lb		0.00078	n	0.00078	0.00015	2	0.0012	2.7
0.003 lb		0.00059	n	0.00059	0.00012	2	0.00099	2.7
0.002 lb		0.00066	n	0.00066	0.00011	2	0.00087	2.7
0.001 lb		0.000182	n	0.000182	0.000083	2	0.0007	2.7
0.001 lb	*	0.000195	n	0.000195	0.000083	2	0.0007	2.7

#### Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

10/9/2017

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

## Calibration Certificate of Mass

**Calibration Date:** October 2, 2017

**Certificate Number:** 2017-027-5

**Submitted By:** FSCP Area 90  
1118 Eldorado Dr.  
Omaha, NE 68514

**Point of Contact:** Brian Heskin  
Ph. 402-496-0843  
**email:** [Brian.heskin@nebraska.gov](mailto:Brian.heskin@nebraska.gov)  
**PO Number:**

**Test Item(s):** 20 lb Kit  
**Serial Number(s):** WM 6D98/17649  
**Manufacture:** Tromner/Rice Lake  
**Condition:** Good (some wear)

**Artifact(s) Description:**

**Date Received:** September 22, 2017

**ID / Asset Number:** N/A  
**Class Specification:** NIST Class F  
**Material:** Stainless Steel

**Reference Standards Used:**

NSL lb standards

**Procedure Used:**

NIST HB 6969, SOP 8

**Metrologist:**

JPL

**Equipment Used:**

Sartorius CCE6 Sartorius CC 1201  
Mettler AT 106 Sartorius CC1000S

**Environmental Cond.** Temp: 22.45 °C Pressure: 768.603 mmHg Relative Humidity: 49 %

**Pertinent Information**

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale verses 8.0 g/cm<sup>3</sup> reference mass density and an air density of 1.2 mg/cm<sup>3</sup> at 20 °C.

**Traceability Statement**

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

**Uncertainty Statement**

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: October 2, 2017

Certificate Number: 2017-027-5


### Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty ± (g)	(k) factor	NIST Class F MPE ± (g)	Assumed Density (g/cm <sup>3</sup> )
10 lb	1	0.174	n	0.174	0.054	2	0.45	7.84
5 lb	2	0.076	n	0.076	0.028	2	0.23	7.84
2 lb	3	0.042	n	0.042	0.011	2	0.091	7.84
2 lb	4	0.034	n	0.034	0.011	2	0.091	7.84
1 lb	5	0.0210	n	0.0210	0.0083	2	0.07	7.84
0.5 lb	6	0.0179	n	0.0179	0.0054	2	0.045	7.84
8 oz	17	0.0106	n	0.0106	0.0054	2	0.045	7.84
4 oz	18	0.0014	n	0.0014	0.0028	2	0.023	7.84
2 oz		-0.0044	n	-0.0044	0.0013	2	0.011	7.84
1 oz		0.00217	n	0.00217	0.00064	2	0.0054	7.84
1/2 oz		0.00099	n	0.00099	0.00034	2	0.0028	7.84
1/4 oz		-0.00029	n	-0.00029	0.00021	2	0.0017	7.84
1/8 oz		-0.00052	n	-0.00052	0.00016	2	0.0013	7.84
1/16 oz		0.00000	n	0.00000	0.00013	2	0.0011	7.84
1/16 oz	*	-0.00052	n	-0.00052	0.00013	2	0.0011	7.84
0.2 lb		0.0084	n	0.0084	0.0022	2	0.018	7.84
0.2 lb	*	0.0088	n	0.0088	0.0022	2	0.018	7.84
0.1 lb		0.0041	n	0.0041	0.0011	2	0.0091	7.84
0.05 lb		0.00175	n	0.00175	0.00054	2	0.0045	7.84
0.02 lb		0.00068	n	0.00068	0.00022	2	0.0018	7.84
0.02 lb	*	0.00035	n	0.00035	0.00022	2	0.0018	7.84
0.01 lb		0.00046	n	0.00046	0.00018	2	0.0015	7.84
0.005 lb		0.00013	n	0.00013	0.00015	2	0.0012	2.7
0.002 lb		0.00000	n	0.00000	0.00011	2	0.00087	2.7
0.002 lb	*	-0.00012	n	-0.00012	0.00011	2	0.00087	2.7
0.001 lb		0.000093	n	0.000093	0.000083	2	0.0007	2.7

#### Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

10/9/2017

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.