

Calibration Date: 8/20/2025

**Certificate of Calibration
of Volume Transfer**

Certificate Number: 2025-139-2

Items Submitted:

Quantity	Nominal Volume	Manufacturer	Type
2	5 gal	Seraphin	Test Measure 4" Neck

Submitted By: FSCP Area 80
3721 West Cuming St.
Lincoln, NE 68524

POC: NDA-Weights and Measures
402-471-3422
agr.wam@nebraska.gov

Test Results

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (1/F)	As Found Volume Delivered @ 60 °F	As left Volume Delivered @ 60 °F	Uncertainty (U)	(k)
5 gal	4393-5-B	304 SS	0.0000288	4.9983 gal	4.9983 gal	0.0012 gal	2.01
5 gal	39423 C	304 SS	0.0000288	5.0009 gal	5.0009 gal	0.0012 gal	2.01

The data in this report only applies to those items specifically listed on this report.

Volume delivered at 60°F after a 30 second pour and 10 second drain for test measures. For provers a 30 second drain time would apply.

Conversion Factors:

1 gal = 231 in³
1 gal = 3.785 412 E-03 m³

Traceability Statement:

The artifact(s) described in this report have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this report is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this report.

Uncertainty Statement:

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors. The combined standard uncertainty is multiplied by a coverage factor (*k*), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken.

Pertinent Information:

The artifact(s) listed above have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and uncertainty exceed 95% of the maximum permissible error. All of the tolerances and specifications were evaluated according to NIST HB 105-3 (2010).

Condition of Item(s) Submitted for Calibration:

Cleaned and ready for calibration

Laboratory Reference Standard Used:

5 gal SP NE 1586

Treatment of Item(s) before Calibration:

Tested as Found

Procedure Used:

NISTIR 7383, SOP 19 (2019)


Environmental conditions at time of calibration:

Temp °C	23.1	Humidity %	43.0
Pressure mmHg	733.60		

Water temperature at time of calibration:

74.39 °F

Date Submitted: 8/18/2025



Joel P. Lavicky, Metrologist

8/25/2025

Issue Date:

This document does not represent or imply endorsement by the State of Nebraska, the Nebraska Standards Laboratory or NIST. This document may not be reproduced, except in full, without the written permission of the Nebraska Standards Laboratory

Calibration Certificate of Mass

Calibration Date: August 19, 2025

Certificate Number: 2025-139-1

Artifact(s) Owner: FSCP Area 80
3721 West Cuming St.
Lincoln, NE 68524

Submitted by: State of Nebraska
Point of Contact: NDA-Weights and Measures
Ph. 402-471-3422
email: agr.wam@nebraska.gov

Test Item(s): 20-cast weights
ID / Asset Number: Area 80
Manufacture: Rice Lake
Material: Cast iron

Artifact(s) Description:

Date Received: August 18, 2025
Serial Number(s): See next page
Class Specification: NIST Class F
Condition: Good (some wear)

Reference Standards Used:

NSL lb standards

Procedure Used:

NIST HB 6969, SOP 8 (2019)

Metrologist:

JPL

Equipment Used:

Mettler XPR32003

Environmental Cond.

Temp: 23.7 °C **Pressure:** 732.3 mmHg **Relative Humidity:** 49.5 %

Pertinent Information

- The artifact(s) listed in this document have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. **RED** print indicates an out-of-compliance reading. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and the uncertainty exceed 95% of the maximum permissible error. All of the tolerances and design specifications (except density, hardness and magnetism) were evaluated according to ASTM E617 (2023) and/or NIST HB 105-1 (2019) for the artifacts designated class.
- All corrections stated in this report correlate to a "Conventional Mass" (CM), also known as "apparent mass", scale versus 8.0 g/cm³ reference mass density and an air density of 1.2 mg/cm³ at 20 °C.
- It is the end user's responsibility to verify that the weights meet the accuracy requirements outlined in NIST Handbook 44 (2022), Appendix A Fundamental Considerations, when using the weights for calibration of commercial (Legal for Trade) scales.

Traceability Statement

The artifact(s) described in this certificate have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this certificate is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this certificate.

Uncertainty Statement

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors associated with air buoyance corrections. The combined standard uncertainty is multiplied by a coverage factor (k), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the *Guide to the Expression of Uncertainty in Measurement (2008, revised 2012)*. Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken. Magnetic testing has not been performed, therefore, there are no components for the effects of it in the uncertainty budget.

Calibration Date: August 19, 2025

Certificate Number: 2025-139-1


Calibration Results

Nominal Mass	Serial Number / ID	As Found Conventional Mass Correction (g)	Adjusted (Y/N)	As Left Conventional Mass Correction (g)	Uncertainty $\pm$ (g)	(k) factor	NIST Class F MPE $\pm$ (g)	Assumed Density (g/cm ³)
25 lb	NE-1	0.01	N	0.01	0.14	2.01	1.1	7.2
25 lb	NE-2	0.53	N	0.53	0.14	2.01	1.1	7.2
25 lb	NE-3	-0.44	N	-0.44	0.14	2.01	1.1	7.2
25 lb	NE-4	0.10	N	0.10	0.14	2.01	1.1	7.2
25 lb	NE-5	-0.70	N	-0.70	0.14	2.01	1.1	7.2
25 lb	NE-6	-0.37	N	-0.37	0.14	2.01	1.1	7.2
25 lb	NE-7	-0.16	N	-0.16	0.14	2.01	1.1	7.2
25 lb	NE-8	-0.11	N	-0.11	0.14	2.01	1.1	7.2
25 lb	NE-9	-0.70	N	-0.70	0.14	2.01	1.1	7.2
25 lb	NE-10	-0.06	N	-0.06	0.14	2.01	1.1	7.2
25 lb	NE-11	-0.32	Y	0.01	0.14	2.01	1.1	7.2
25 lb	NE-12	-0.25	N	-0.25	0.14	2.01	1.1	7.2
25 lb	NE-13	-0.23	N	-0.23	0.14	2.01	1.1	7.2
25 lb	NE-14	-0.16	N	-0.16	0.14	2.01	1.1	7.2
25 lb	NE-15	-0.27	N	-0.27	0.14	2.01	1.1	7.2
25 lb	NE-16	-0.64	N	-0.64	0.14	2.01	1.1	7.2
25 lb	NE-17	-0.54	N	-0.54	0.14	2.01	1.1	7.2
25 lb	NE-18	-0.20	N	-0.20	0.14	2.01	1.1	7.2
25 lb	NE-19	-0.21	N	-0.21	0.14	2.01	1.1	7.2
25 lb	NE-20	0.04	N	0.04	0.14	2.01	1.1	7.2

Conversion Factors

1 ounce (avoirdupois) (oz) = 28.349 52 g

1 pound (avoirdupois) (lb) = 453.592 37 g exactly



Joel P. Lavicky Metrologist

8/25/2025

Date of Issue

The results in this certificate only applies to those item specifically listed in this certificate. This certificate cannot be considered complete unless it contains all pages. This document may not be reproduced except in full, without the written consent of the Nebraska Standards Laboratory.

Calibration Date: 8/20/2025

**Certificate of Calibration
of Volume Transfer**

Certificate Number: 2025-139-3

Items Submitted:

Quantity	Nominal Volume	Manufacturer	Type
3	5 gal	Seraphin	"Special" J Prover

Submitted By: FSCP Area 80
3721 West Cuming St.
Lincoln, NE 68524

POC: NDA-Weights and Measures
402-471-3422
agr.wam@nebraska.gov

Test Results

Nominal Volume	Serial Number	Material	Cubical Coefficient of Expansion (1/°F)	As Found Volume Delivered @ 60 °F	As left Volume Delivered @ 60 °F	Uncertainty (U)	(k)
5 gal	20-64572-10	304 SS	0.0000288	5.00174 gal	4.99980 gal	0.00087 gal	2.01
5 gal	20-64572-11	304 SS	0.0000288	5.00072 gal	5.00072 gal	0.00087 gal	2.01
5 gal	20-64572-12	304 SS	0.0000288	5.00042 gal	5.00042 gal	0.00087 gal	2.01

The data in this report only applies to those items specifically listed on this report.

Volume delivered at 60°F after a 30 second pour and 10 second drain for test measures. For provers a 30 second drain time would apply.

Conversion Factors:

1 gal = 231 in³
1 gal = 3.785 412 E-03 m³

Traceability Statement:

The artifact(s) described in this report have been compared to the Standards of the State of Nebraska. The Standards of the State of Nebraska are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The calibration number for this report is the only unique calibration number to be used in referencing measurement traceability for the artifact(s) described in this report.

Uncertainty Statement:

The combined standard uncertainty includes uncertainties reported for the standard, uncertainties associated with the measurement process, uncertainties for any observed deviations from reference values which are less than surveillance limits and the standard uncertainty for any uncorrected errors. The combined standard uncertainty is multiplied by a coverage factor (k), to give the expanded uncertainty, which defines an interval with a 95.45 percent level of confidence. The expanded uncertainty presented in this report is consistent with the Guide to the Expression of Uncertainty in Measurement (2008, revised 2012). Some components of the calibration can be evaluated through a Type A evaluation, or the method of evaluation of uncertainty by the statistical analysis (standard deviation) from the observations taken.

Pertinent Information:

The artifact(s) listed above have been found and/or left within the maximum permissible error for the specification stated above, except as noted. An artifact is considered in-compliance when the correction plus the measurement uncertainty is equal to or less than the maximum permissible error. It is the decision of the Laboratory to adjust the artifact(s) when the sum of the correction and uncertainty exceed 95% of the maximum permissible error. All of the tolerances and specifications were evaluated according to NIST HB 105-3 (2010).

Condition of Item(s) Submitted for Calibration:

Cleaned and ready for calibration

Laboratory Reference Standard Used:

5 gal SP NE 1586

Treatment of Item(s) before Calibration:

Tested as Found

Procedure Used:

NISTIR 7383, SOP 19 (2019)


Environmental conditions at time of calibration:

Temp °C	23.3	Humidity %	45.0
Pressure mmHg	733.30		

Water temperature at time of calibration:

74.59 °F

Date Submitted: 8/18/2025



Joel P. Lavicky, Metrologist

8/25/2025

Issue Date:

This document does not represent or imply endorsement by the State of Nebraska, the Nebraska Standards Laboratory or NIST. This document may not be reproduced, except in full, without the written permission of the Nebraska Standards Laboratory